NASA Contractor Report 189172

Cryogenic On-Orbit Liquid Depot Storage Acquisition and Transfer (COLD–SAT) Experiment Subsystem Instrumentation and Wire Harness Design Report

Lawrence G. Edwards Analex Corporation Brook Park, Ohio

March 1994

(NASA-CR-189172) CRYOGENIC N94-29405
ON-ORBIT LIQUID DEPOT STORAGE
ACQUISITION AND TRANSFER (COLD-SAT)
EXPERIMENT SUBSYSTEM Unclas
INSTRUMENTATION AND WIRE HARNESS
DESIGN REPORT Final Report (Analex
Corp.) 72 p G3/18 0003794

Prepared for Lewis Research Center Under Contract NAS3-25776



~

CRYOGENIC ON-ORBIT LIQUID DEPOT STORAGE ACQUISITION AND TRANSFER (COLD-SAT) EXPERIMENT SUBSYSTEM INSTRUMENTATION AND WIRE HARNESS DESIGN REPORT

Lawrence G. Edwards Analex Corporation Brook Park, Ohio

SUMMARY

This report describes an instrumentation and wire harness design for the COLD-SAT experiment subsystem. The design incorporates transducers, signal conditioning systems, and wire harness components that were chosen, whenever possible, based on their past successful use with liquid hydrogen (LH₂) and on previous spaceflight systems. Electrical current excitation levels and data acquisition ranging were designed to meet the temperature measurement requirements using space-flight qualified platinum resistance thermometers with existing 8-bit space-flight qualified data acquisition systems. Cone shaped LH₂ temperature and liquid/vapor sensors were recommended to minimize false measurements in low-gravity conditions. The need for a 12-bit data acquisition system to provide improve resolution and accuracy for critical measurements is addressed, and an experimental data unit design is presented. Error analyses were performed on all instrument candidates, and the influence of physical parameters such as temperature and pressure on the overall measurement accuracies was determined. Most experimenters' measurement requirements can be met by the design; however, some measurement techniques and instrumentation need development. Two-phase flow detection, leak detection systems, and mass gauging are areas that need development. Heat conduction to the LH2 tanks was recognized as a chief concern early in the design effort. Wire materials, wire multiplexing techniques, and cryogenic operable pressure transducers were selected to minimize this problem.

A list identifying all measurements and the transducers to be used to make the measurements is presented. Schematic drawings showing the locations of the transducers are presented, and a total system wire harness was designed to show the feasibility and hardware requirements of the system.

INTRODUCTION

Acquiring the technology to store and transfer cryogens in space is mandatory for future long-duration space flight missions. COLD-SAT, an acronym for cryogenic orbiting liquid depot for storage, acquisition, and transfer, is a spacecraft-experiment system designed to obtain this technology. The spacecraft will contain a LH₂ supply tank, a large receiver tank, and a small receiver tank. Experiments investigating LH₂ tank pressure control, receiver tank chilldown methods, and receiver tank fill techniques will be conducted. The success of the COLD-SAT experiments will be directly related to the measurements made and the accuracy with which the measurement data are acquired and processed. The instrumentation, signal conditioning, and measurement techniques used to make the various measurements is critical in determining the success of the mission.

The instrumentation must be capable of surviving the stresses of launch and the environment of space and be reliable over the course of the COLD-SAT mission. The instrumentation must not interfere with or distort the processes it measures. Transducers must operate in liquid and gaseous hydrogen environments. Electrical wiring materials and techniques must minimize heat conduction to the cryogenic tanks.

This report presents a detailed design for the instrumentation, signal conditioning, and wire harness. In developing the design, instrumentation candidates capable of meeting the measurement requirements were compared, and recommendations were proposed based on error analysis, reliability, and performance histories. The location, mounting techniques, and electrical harnessing of the instruments are addressed. Experiment signal conditioners and telemetry interface systems that are designed to maximize the return of experimental data while minimizing interference with experiment phenomena are presented. System component weights and power requirements were estimated. Specific instrumentation models are not identified in this report to avoid the appearance of NASA or Analex Corp. endorsement of commercial products.

This report is meant to be used in conjunction with the overall COLD-SAT Technical Memorandum report which contains detailed descriptions of the spacecraft, experiment requirements, and mission philosophy.

This report documents the NASA Lewis Research Center in-house COLD-SAT instrumentation and wire harness design effort conducted under Contract NAS3-25776. The feasibility design was completed for presentation at a nonadvocate review of the total COLD-SAT program held in June 1990. The findings of the review committee were that the project was feasible. The program, however, was found unacceptable due to cost.

ACKNOWLEDGMENTS

The author is grateful to all the members of the COLD-SAT design team, NASA Lewis Research Center's Cryogenic Fluid Technology Office, Analex Corp., editors, and the numerous others who have contributed to the completion of the design effort.

COLD-SAT EXPERIMENT INSTRUMENTATION AND ELECTRONICS SUBSYSTEM TT&C

Interface Definition

The experiment system instrumentation and electronics subsystem consists of the transducers required to convert physical process parameters, such as temperature, pressure, flow rate, and acceleration, into electrical equivalent signals. The transducers require electrical excitation signals, which are provided by power supplies contained in the signal-conditioning electronic boxes. The output signal of the transducers must be conditioned to a form compatible for input to the spacecraft telemetry tracking and command (TT&C) subsystem. This is accomplished by amplifiers, filters, and analog-to-digital signal converters. The experiment system contains three high-resolution, 12-bit analog-to-digital converter signal conditioning units, called experiment data units, or EDU's. The experiment system instrumentation and electronics subsystem also contains the instrumentation and electrical component wire harnesses, sensor mounting hardware, heaters, liquid-level capacitive probe signal conditioning unit, accelerometer, and fluid mixer power unit. Table 1 lists the experiment system electronic boxes with their estimated weights and power requirements. Figure 1 is an artist's concept of the COLD-SAT spacecraft that shows the location of the electronic bays.

The experiment electrical and electronic boxes receive power from the spacecraft electrical system. Heaters, valves, mixer motor, and most instrumentation receive electrical power through the TT&C system. The TT&C system contains a command telemetry unit (CTU) and two remote command telemetry units (RCTU's), which provide transducer electrical power, sensor output signal conditioning, and 8-bit analog-to-digital signal conversion. The TT&C system also contains relay sequencer units (SEQ's), which distribute electrical power to the experiment subsystem valves and heaters. Figure 2 is a block diagram of the COLD-SAT experiment system and its interface with the TT&C system. The total number of sensors and the number of sensors assigned to each signal conditioner is shown.

Transducers

Transducers consist of the sensors and built-in compensation networks required to convert physical properties into equivalent electrical signals. Temperature, pressures, flow rates, acceleration, liquid level, vapor detection, mixer motor speed, power, and valve status transducers are required for monitoring the physical properties and status of the COLD-SAT experiment fluids and systems.

Data Acquisition Systems

Figure 3 is a simplified block diagram representation of a COLD-SAT experiment data acquisition system (DAS). The DAS acquires the analog output signals of the transducers, conditions the signals by means of amplifiers, filters, and translators and converts the signal to an equivalent digital form required by the TT&C system.

To minimize circuitry, electronic multiplexers are used to switch the transducer output to a single instrumentation amplifier stage where the signal is converted to a 0- to 5-V level. The voltage output of the amplifier stage is converted to an equivalent digital form for the telemetry system by analog-to-digital converters. The convertors are defined by the number of binary digits (bits) they use to represent an analog signal range. The resolution, or smallest detectable change, of the DAS is equal to the span of the measurement system divided by the decimal equivalent of the number of bits. COLD—SAT experiment measurements will be monitored with two types of DAS systems: 8-bit analog-to-digital systems, located in the TT&C telemetry units, and 12-bit analog-to-digital systems, designated as experimental data units (EDU's).

Telemetry tracking and command 8-bit data acquisition systems.—The TT&C system provides most of the signal conditioning electronics for the COLD-SAT instrumentation. The TT&C system uses 8-bit analog-to-digital converters located in the systems command and remote command telemetry units (RCTU's). The 8-bit converters set the measurement resolution to 1 part in 256 of the analog measurement range. The maximum error between the analog input and its digitized equivalent can be as great as 1/2 the resolution value, or ±0.195 percent of the full-scale range. This error component is known as the quantization error (ref. 1). The system's multiplexers have leakage current components that produce measurement errors due to the voltage drop they produced across resistive components in the circuit. Each signal conditioning card contains eight multiplexers that can result in a total leakage current component eight times greater than the specified value of one multiplexer. The uncertainties of the signal conditioning electronics were included in the measurement error analysis and influenced the selection of a sensor type. Table 2 lists the estimated root sum square (rss) signal conditioning uncertainties for the anticipated sensor measurement ranges. The values were based on existing 8-bit flight qualified data acquisition system capabilities. The rss method of combining uncertainties was used for all measurement error analysis because of the unlikely probability that all uncertainties are coherent (ref. 2).

Experimental data units.—Three EDU's were designed with 12-bit analog-to-digital converters to provide improved accuracy and resolution and to meet measurement requirements incapable of being met by the 8-bit systems. The EDU design and 8-bit signal conditioning system designs were identical except for the use of a 12-bit analog-to-digital converter and only one multiplexer per circuit card. The EDU's would provide excitation and signal conditioning electronics for the flow-measurement systems, accelerometer, and high-resolution tank-pressure measurement transducers. Figure 4 is a simplified block diagram of an EDU measurement system. Figure 5 shows the major signal conditioning errors of the two types of DAS units estimated for the 30-mV measurement range. The use of a 12-bit analog-to-digital converter and only one multiplexer per card reduced the quantization error to 0.012 percent of the range and reduced the leakage current influence by 1/8th. The qualification of a 12-bit analog-to-digital converter for space flight is required for this system to be feasible.

Harnessing

The experiment subsystem instrumentation and electrical power component harness consists of the electrical wiring and connectors required to connect the transducers and electrical powered components to their associated signal conditioners, power supplies, and relay boxes.

Most of the instrumentation and electric powered components will be thermally bonded to the LH₂ tanks. The temperature gradient experienced by the component wiring will result in a large thermal energy input into the tanks and an excessive boiloff loss of hydrogen if copper wire were used. To minimize this problem, unconventional wire material will be used to electrically connect the cryogenic tank instrumentation and hardware to the warm copper wire harness of the spacecraft wire trays. Manganin wire will be used for instrumentation wiring. Phosphor bronze wiring will be used for valves and heater applications.

Manganin has a distinct advantage over other wire materials for cryogenic use because of its very low thermal conductivity at low temperatures (ref. 3). Unfortunately, being a poor thermal energy conductor also results in manganin being a poor electrical conductor. However, the current requirements for most of the instrumentation is below 15 mA, and 24 American wire gauge (AWG) diameter manganin wire is adequate. Four-wire measurements will be made on temperature sensors to eliminate the wire voltage drop of the sensors' current carrying leads from the voltage-measurement circuit. The constant current operation of pressure transducers will eliminate wire voltage-drop influence from this measurement.

Phosphor bronze wire provides a combination of fairly low thermal conductivity with reasonable resistivity and was selected for valves, motors, and heaters that have current ratings that would require excessively large-diameter manganin wire (ref. 4). The use of phosphor bronze wire would reduce the electrical components harness weight and limit heat conduction to a tolerable level.

Connectors and receptacles used for cryogenic tank wire feedthroughs will be similar to the type used on D-1A Centaur launch vehicles. Bulkhead connector assemblies exposed to temperatures below – 55 °F will conform to George C. Marshall Space Flight Center specification 40M38294. Connectors exposed to temperatures above –55 °F will be series MIL–C-38999.

MEASUREMENT REQUIREMENTS

The COLD-SAT experiment matrix consists of eight primary experiments. From the primary experiment data requirements the following measurements were identified to be necessary:

- · High-resolution LH₂ and GH₂ temperatures (type A)
- Tank structure temperatures (type AB)
- · Hydrogen liquid-vapor level
- · Liquid-vapor two-phase flow detection
- · Tank pressures and flow device pressure drops
- · Liquid-hydrogen flow rates
- · Mixer flow rates
- · Thermodynamic vent systems (TVS) and vent flow rates
- · Spacecraft acceleration
- · Valve status indication
- · Electrical power supplied to heaters and mixer motors.

Transducers were selected for each type of measurement based on measurement range, accuracy, and reliability. Signal conditioning techniques were developed to achieve best possible measurement accuracy based on a measurement error analysis. Table 3 lists the required measurements, the measurement range, transducer type, and numbers. Calculated measurement uncertainties and desired sampling rates are also listed. A detailed measurement list, which provides traceability of each measurement to its primary experiment requirement document, is located in the appendix.

Temperature Measurement Requirement

Two primary temperature measurement requirements were determined from the experiment requirements.

High-resolution LH_2 and GH_2 temperatures (type A).—The capability of measuring and resolving small temperature gradients existing from the LH_2 tank walls into the bulk of the fluid and also at the LH_2 — GH_2 interface is required. Temperature sensors capable of fulfilling this requirement would provide temperature measurements of structure and fluid in the saturated and subcooled hydrogen temperature ranges. An uncertainty less than or equal to ± 0.2 °R for the 20 to 50 °R temperature range is required for these measurements. This high-accuracy temperature measurement was designated as type A.

The saturation temperature of hydrogen varies from 30.8 °R at 5 psia to 45.4 °R at 50 psia (ref. 5). This pressure range covers the operating range of the cryogenic tanks and thermodynamic vent systems. The temperature measurement accuracy over this range was considered critical.

Tank structure temperatures.—Temperature measurements of tanks, structure components, and fluids during chilldown experiments and periodically during the course of the COLD-SAT mission must be measured. A total temperature range measurement from 36 to 540 °R is required. This total temperature range measurement was designated type AB.

Temperature Sensor Candidates

A number of temperature transducer types exist that could be used to measure the required temperature ranges (ref. 6). Three popular thermoresistive temperature sensors used for cryogenic temperature measurements are germanium resistance thermometers (GRT), platinum resistance thermometers (PRT), and silicon diodes. Thermoresistive temperature measurement is based on the change in a metal or semiconductor impedance as a function of temperature.

Other temperature measurement sensors include thermocouples and thermistors. Thermocouples were not considered because of their low sensitivities and poor accuracies. Thermistors were not considered because of their rapidly changing sensitivities and the need for multiple range-switching signal-conditioning systems.

Platinum resistance thermometers.—Platinum resistance thermometers (PRT's) belong to the resistance temperature detector (RTD) family. Resistance temperature detection is based on the increase in resistivity of a metal conductor with increase in temperature. Ideally, the metal used for the temperature sensor would be pure and would be mounted in a strain-free manner. The resistance of the sensor, then, would be dependant only on temperature and metal geometry. The basic relation between sensor resistance, geometry, and temperature is

$$R_t = \sigma L/A = R_o \left[1 + \alpha \left(T - T_o \right) \right] \tag{1}$$

where A is the area, L is the length, R_t is the resistance, R_o is the ice point resistance, T is the temperature, T_o is the ice point temperature, T_o is the resistivity, T_o is the temperature coefficient of resistance.

Platinum resistance thermometers are classified according to their water ice point resistance (R_o) values and can range in value from tens to thousands of ohms. They have been used extensively in aerospace and cryogenic applications and have also been mounted and used in special cone-shaped probes designed for use in low-gravity environments to minimize false ullage temperature readings (ref. 7).

Germanium resistance thermometers.—Germanium resistance thermometers (GRT's) typically consist of a specially grown and doped germanium crystal mounted in a strain-free manner within a copper metal can. Four wire leads of very-small-gauge copper or low-thermal-conductivity wire are attached to the sensor to provide electrical excitation and output signal sensing connections. Germanium resistance

thermometers have negative resistivity temperature coefficients, and their values of resistance and sensitivity can vary by several orders of magnitude over the temperatures from 18 to 180 °R. To prevent self-heating errors, manufacturers recommend that the sensor be excited by a power source such that the potential across the sensor is kept at 10 mV or less. Measurements taken with GRT's would be limited to type A. A different type of temperature sensor would be required to monitor the warmer temperature ranges.

Silicon diode temperature sensors.—Silicon diode temperature sensors are nonlinear semiconductor devices that exhibit negative temperature coefficients of resistivity. They exhibit very high sensitivity (millivolts per degree) at cryogenic temperatures and good sensitivity above the cryogenic range. The sensors are small and rugged, and their high impedance may allow two-wire measurements to be made for less accurate measurement requirements. Silicon diodes are attractive as total range temperature sensors that could monitor the chilldown rate of the COLD-SAT tank structures. The low 10-µA excitation recommended for silicon diode operation virtually eliminates sensor self-heating errors.

Temperature Measurement Error Sources

Temperature sensor calibration inaccuracy.—Calibration of the temperature sensors is required to accurately characterize the specific sensor's resistance versus temperature characteristic and to obtain measurement uncertainties of ≤ 0.2 °R. Inaccuracies in the calibration process are due to temperature standard, calibration media, and measurement equipment errors. Commercial temperature sensor calibration uncertainties are typically less than ± 0.05 °R.

Interpolation errors.—The voltage output response of the calibrated temperature sensor is converted to its temperature equivalent by use of a polynomial equation. The accuracy of the temperature interpolation equation is dependant on the calibration measurement accuracy and the number of calibration points. Interpolation error can also occur when linear interpolation is used between calibrated points. Interpolation errors of less than ± 0.04 °R can be obtained by proper specification of the number of calibration points and the temperature increment size on the calibration table.

Data acquisition system error.—The largest uncertainty in the COLD-SAT temperature measurement is due to data acquisition system limitations. The DAS voltage measurement error for the temperature sensor candidates was estimated using the DAS voltage measurement range errors listed in table 2. Table 4 lists the estimated excitation signals, output signal levels, and data acquisition inaccuracies for the sensor candidates.

A concern exists regarding the ability to flight qualify a $10-\mu A$ constant current source required for the silicon diode sensors. However, for comparison purposes a $10-\mu A$ flight-qualified current source was considered feasible with an excitation error of less than ± 0.1 percent.

To prevent self-heating errors, the recommended excitation level for GRT sensors is a current that produces a potential drop across the sensor of 10 mV or less. An electronic power supply that would automatically regulate the sensor current so that the sensor potential remains at 10 mV would be required. The excitation current would be detected across a precise resistor and temperature would be calculated from the current level magnitude.

Platinum resistance thermometer temperature measurement design and error analysis.—The PRT sensitivity to meet the high-accuracy temperature measurement requirement of ± 0.2 °R must be greater than the ratio of the voltage measurement range error divided by the inaccuracy requirement of 0.2 °R; that is, assuming a rss error for the 125-mV measurement of 0.36 percent or 0.45 mV:

Sensitivity =
$$mV/^{\circ}R \ge (0.45 \text{ mV}/0.2 ^{\circ}R) \ge 2.25 \text{ mV}/^{\circ}R$$
 (2)

The sensitivity of commercially available ice-point resistance PRT's was calculated as a function of excitation current magnitude. Table 5 lists the typical resistance-temperature relationship for a $1000-\Omega$ PRT.

The change in PRT resistance as a function of temperature was calculated. The product of this value and the excitation current magnitude of 10 mA yields the sensitivity of the sensor at that temperature. The temperature equivalent of the DAS error was calculated by taking the product of the inverse of the sensitivity factor and the millivolt DAS error value. Figure 6 shows that the 1000- Ω (R_o) PRT, when excited by a 10-mA constant current, could meet the sensitivity requirement with a DAS uncertainty of ± 0.2 °R over the temperature range of 29 to 50 °R. The total rss combination of sensor calibration, interpolation, and DAS error would still be less than ± 0.21 at 29 °R.

Self-heating error considerations.—A major concern was the magnitude of error that could occur due to self-heating of the PRT sensor by the 10-mA excitation current. Self-heating causes the sensor's temperature to rise to a value greater than the temperature of the environment it is to measure. To minimize self-heating, the sensors would be energized for short duty cycles of approximately 1.0 percent. The effective power dissipated by the sensor would be 1×10^{-4} that of its continuous operation duty cycle value.

Total range PRT temperature measurement (type AB) error analysis.—The TT&C DAS uses programmable gain amplifiers that will allow dual measurement ranging. This capability allows a large temperature span to be monitored at a reduced accuracy. Error analysis performed on PRT sensors showed that a $1000 R_0$ PRT excited by a 1-mA constant current source could measure the temperature range from 29 to $560 \,^{\circ}$ R with an 8-bit DAS uncertainty within $\pm 2.0 \,^{\circ}$ R. A DAS range change from the $125 \,^{\circ}$ mV range to the $1.25 \,^{\circ}$ V range would occur at temperatures above $110 \,^{\circ}$ R. Figure 7 shows the results of this error analysis.

Silicon diode, 8-bit DAS temperature measurement error analysis.—The estimated temperature measurement error for a commercially available silicon diode was determined using the manufacturer's diode sensitivity values and the estimated 8-bit DAS, 1.25-V range measurement error. Below 33 °R the diodes output voltage would exceed the 1.25-V range limit, and a DAS range change would be required. The temperature measurement error results for the 1.25-V range are shown in figures 5 and 6.

High-accuracy liquid hydrogen temperature range measurement.—The fraction of a volt deviation in the diode output that occurs in the critical LH₂ temperature range rides on a 1-V level. To maximize the accuracy and resolution in this critical measurement range a -1.13-V offset could be algebraically summed with the diodes output. This would result in the shifting of the output voltage so that the more sensitive 125-mV measurement range could be used over the range from 33 to 43 °R. An amplifier range change would be required for temperatures below 33 °R and above 43 °R. Figure 8 shows the results of this analysis. The performance of the type A PRT measurement is also shown for comparison purposes.

GRT - 8-bit temperature measurement design and error analysis.—The contribution of data acquisition error on temperature measurement error for a GRT system was estimated using the values of sensitivity for a commercially available sensor. The sensor output voltage would be maintained at a 10-mV level, and the resultant current level would be determined from the voltage drop measured across a sensing resistor. The limitation of this measurement system is the accuracy of measuring and maintaining the 10-mV GRT voltage drop.

The estimated 10-mV measurement range error for the 8-bit system is ± 0.18 mV. The temperature error analysis results (table 6) show that the temperature equivalent of the voltage uncertainty would exceed ± 0.27 at 36 °R. The results are also shown in figure 5 for comparison purposes.

Wire voltage drop errors.—Manganin wire was selected for instrumentation hookup wiring because of its low thermal conductivity. Manganin, however, has a greater resistivity than copper, and the voltage drop across the sensor wiring would be a significant source of error. To eliminate this problem, the temperature sensors will be used in four-wire measurement circuits: Two wires supply power to the sensor, and two wires couple the sensor voltage drop to the high-impedance voltage measurement circuit of the DAS. Virtually no current flows in the sensing lead; thus the excitation current (I) wire voltage drop $(I \times R)$ is eliminated from the voltage measurement. Silicon diodes, because of their very high impedance and low excitation current level, may be used in two-wire systems for low-accuracy measurements.

Thermoelectric potential effects.—Thermoelectric potential effects due to temperature gradients across the instrumentation, wiring inhomogeneities, and dissimilar metal junction temperatures may occur. To compensate for this problem, the measurement system would be programmed to periodically take a sensor measurement without current excitation. This would determine the magnitude of the thermocouple effects and allow the offset corrections to be made for high-accuracy measurement requirements.

Low-gravity influence on measurement error.—Under low-gravity conditions, surface tension forces dominate, and liquid films could wick over temperature sensors in an otherwise gaseous environment resulting in measurement errors. Special cone shaped sensors designed to wick liquid films away from the sensor tip have been designed to minimize this problem (ref. 7).

Temperature Sensor Selection

Platinum resistance thermometers were recommended for COLD-SAT because of their successful use in previous aerospace applications and because of the feasibility of highly accurate measurements with minimal change to existing flight-qualified data acquisition measurement systems. Desired specifications for the temperature sensors are detailed in this section.

High-accuracy type A tank internal fluid temperatures.—A PRT probe similar to that shown in figure 9 is recommended for immersion temperature measurements. The probe would consist of a $1000-\Omega$ (R_o) platinum sensing element mounted in a stainless-steel cone. The cone shape is designed to wick away liquid films from the sensor tip, thereby minimizing false ullage temperature readings under low-gravity conditions. The probe design is very rugged, and the threaded mounting attachment ensures a reliable mount to an instrumentation rake. The probe would be connected with 24-gauge manganin wire to minimize heat conduction. Probes of similar design have been spaceflight qualified by a commercial manufacturer. A simplified schematic of the four-wire PRT excitation and voltage measurement circuit is included in figure 9.

Hydrogen saturation temperature based on absolute pressure.—The hydrogen saturation temperature within the cryogenic tanks can be determined from the tank absolute pressure measurement. Figure 10 shows hydrogen saturation temperature versus saturation pressure (ref. 5). A forth-order polynomial was found to fit this graph with a regression factor of 1.0. The uncertainty of inferring temperature (UT) from the absolute pressure measurement was calculated by taking the product of the polynomial slope and the uncertainty of the pressure measurement (UP) at the pressure (P) of interest. Table 7 lists the results of this analysis using the estimated values of pressure measurement uncertainty for a 50 psia range transducer (table 8). This measurement is valid only for fluid and surface measurements at the saturation temperature. The uncertainty in this requirement necessitates the use of multiple temperature sensors located throughout the cryogenic system.

High-accuracy type A surface temperature measurements.—Platinum resistance thermometers are recommended for the surface measurements because they are capable of providing total measurement inaccuracies of less than ± 0.2 °R over the critical LH₂ range from 29 to 46 °R while using an 8-bit data acquisition system.

Type AB total temperature range measurement.—Platinum resistance thermometers are recommended to perform the surface temperature measurements over the range from 36 to 540 °R. Total temperature range measurements with inaccuracy less than ± 2 °R are feasible with the dual range measurement capability of the data acquisition system.

Mounting considerations.—Surface temperature sensors must be attached to the structure in a manner such that intimate thermal contact is obtained with adequate strength to withstand thermal cycling and stresses of launch. Sensors can be mounted with adhesives or mounted in a high-thermal-conductivity holder that can be attached with screws or spot welded to the structure. The sensor leads should be thermally bonded to the structure so that they are at the same surface temperature. Reference reports detailing recommended mounting techniques are available from the various sensor manufacturers.

Absolute Pressure Measurement Requirements

The absolute pressure of the cryogenic tanks, plumbing systems, and gas pressure supplies must be monitored to determine experiment pressurization rates, system operational status, and fluid thermodynamic state information. Measurement resolution of better than 0.2 psia are desired over the 50-psia operation range of the cryogenic tanks.

Absolute pressure measurements of the cryogenic tanks present unique problems. Standard temperature range transducers could be used by tapping into the system using stainless steel tubes. However, this procedure could result in thermal acoustic oscillations and conduct large amounts of heat energy into the LH₂ tanks. A transducer capable of operating at the LH₂ temperature range that could be directly installed at the cryogenic tank is desired.

Absolute cryogenic rated pressure measurement transducer.—A commercial strain gauge absolute pressure transducer is available that has a compensated cryogenic temperature range 36 to 186 °R. Table 8 lists the specifications for this cryogenic absolute pressure transducer. The transducer can also be calibrated for warmer temperature range operation. This transducer type has been flight qualified for the space shuttle.

Absolute pressure sensor specifications and error analysis.—The major sources of pressure measurement error include transducer inaccuracy, temperature effects, data acquisition error, and wiring errors. Transducer inaccuracy consists of the root sum square combination of nonlinearity, hysteresis, and nonrepeatability. This specification is listed as a percent of full-scale range. The transducers also have a specified compensated temperature range in which their performance change is defined as a function of temperature. The influences of temperature on the transducer zero point and slope are expressed in the thermal zero shift and thermal sensitivity shift specifications.

The pressure transducers used on the COLD-SAT experiment system will experience periods of temperature cycling. The transducers must be calibrated at various temperatures, and the influence of temperature effects on zero and sensitivity shifts determined.

Excitation and wiring errors.—Manganin wiring is recommended to minimize the heat energy conducted into the cryogenic tanks. Manganin wire of 24 AWG size would have an approximate resistance of $0.7~\Omega/\mathrm{ft}$. The voltage drop across the manganin wire by the current level supplied by the 10-V power supply to the $1000\text{-}\Omega$ bridge resistance pressure transducer would be sufficient to change the voltage actually supplied to the transducer. A second voltage measurement directly at the transducer input terminals must be made to determine the actual excitation voltage of the transducer. To eliminate the wire voltage drop problem, the transducers should be calibrated and excited by a equivalent current source equal to the recommended excitation voltage (10 V dc) divided by the transducer bridge resistance.

Data acquisition error.—The full-scale output of the pressure sensor candidates is 30 mV. The estimated 8-bit DAS error for this measurement range is ± 0.67 percent of full scale with a resolution of 0.39 percent of full scale. This system would result in a DAS error of ± 0.365 psia for a 50-psia range measurement with a resolution of 0.196 psia.

High-accuracy pressure measurements will be made with the EDU systems. The EDU system will decrease the DAS error to ± 0.36 percent of full scale and improve the resolution to 0.024 percent of full scale. Table 9 lists the resolution values and psia equivalent errors of the rss combination of transducer inaccuracy (0.25 percent of full scale) and DAS inaccuracy (percent of full scale) for the two types of measurement systems as a function of full-scale pressure range.

Discrete-Point Liquid-Vapor Detection

Liquid-level detectors are required to determine tank liquid level during ground fill operation. Inflight uses include liquid level gauging during acceleration, two-phase flow detection, and low-gravity liquid-vapor distribution profiles of tank fluid.

Discrete-point liquid-vapor detector candidates include thermoresistive elements such as thermistors, carbon resistors, and wire elements. The detectors are excited at an electrical power level at which self-heating of the sensor occurs. Self-heating causes sensor resistance to be significantly different when immersed in liquid from its resistance when immersed in vapor. The sensor resistance change produces a voltage level change that is fed to a comparator circuit. The electronic comparator circuit produces a high- or low-voltage output dependant on the fluid state. The sensor can be used in a bridge circuit configuration and excited by a constant voltage source or singularly excited by a constant current source.

Other methods of determining fluid level include capacitance probes and resistive tape strips. Using these type of detectors with high-wetting fluids under low-gravity conditions could result in inaccurate measurements due to fluid wicking. Discrete-point level sensors could also give erroneous indications of liquid presence under low-gravity conditions due to the presence of a liquid film existing over the sensor. Cone shaped sensors that use the surface tension force experienced by the fluid to wick away the fluid film from the sensor tip are recommended for low-gravity applications.

Recommended sensor.—Cone shaped liquid-vapor discrete-point detectors similar in shape to the temperature probe shown in figure 9 have been used on Centaur upper stage vehicles. A sensor similar to that design containing a thermistor sensing tip is proposed. The probe would be supplied with 24 AWG manganin wire leads to reduce conducted heat flow.

Ground fill liquid hydrogen gauging.—Ground test results indicate that inaccurate discrete-point detector liquid-level readings can occur during tank fill operations because of the vigorous boiling of the hydrogen. Commercially available capacitance probes, however, have performed satisfactorily. Capacitance probes are used for LH₂ gauging for the space shuttle power reactant storage assembly. A capacitance probe that will monitor the fill level from 85 to 100 percent is required for COLD-SAT.

Two-phase flow detection.—Two-phase flow detection inside the liquid acquisition device (LAD) channels and at the vent line exits is required. Thermistors that have very high-sensitivities could be mounted inside the channels or at the entrances to the plumbing lines. The sensors would be excited at a power level where sensor self-heating occurs in vapor but not in liquid. Testing and development of sensors, mounting and signal conditioning is needed to optimize two-phase flow detection.

Flowmeters used on the LH₂ and vent lines should also give two-phase flow indications. Two-phase flow through a turbine meter is to be avoided and will cause noticeable speed increases. Two-phase flow through a differential pressure flow metering device should also be detectable by pressure surges.

Liquid Hydrogen Flow-Rate Measurement

Liquid hydrogen flow rates of 50, 100, and 200 lbm/hr are required for the tank chilldown and LH₂ transfer experiments. Fluid inventories must also be maintained by integrating the LH₂ flow rates over the COLD-SAT experiment time line. Studies performed by independent testing facilities indicate the older flow-measurement techniques using turbine meters or pressure head meters (venturi/orifice) to be the most reliable (ref. 8).

The desired flow rates will be obtained by sizing orifices or venturies to provide the required pressure drops. The pressure drop across a flow-control device is ideally proportional to the flow rate squared and can be monitored with a differential pressure transducer for determining the fluid volumetric flow rate. The use of differential pressure transducers mounted to tubes that are tapped to the flow lines would conduct undesirable levels of heat into the LH₂ tanks and present thermal acoustic oscillation problems. A differential pressure transducer capable of operating in LH₂ is required.

Turbine flowmeters have also been used successfully on LH₂ ground test facilities. Over speeding of the turbine meter by liquid flashing into a vapor is a concern. Target meters in which the flowing liquid imparts a force on the target proportional to the flow rate squared were also considered.

Liquid hydrogen flow-measurement candidate error analysis.—Differential pressure and turbine flowmeters measure volumetric flow rate. Therefore, to determine the mass flow rate, the fluid density

must be known. The LH₂ density will be determined by measuring the fluid's temperature and pressure at the flowmeter inlet. The LH₂ density will be determined from thermal dynamic property tables. The LH₂ mass flow rate error is a function of the volumetric flow-measurement error and the error in density determination.

Estimated LH_2 density error.—The influence of temperature and pressure on LH_2 density was estimated by determining best fit equations that mathematically described the change in LH_2 density as a function of temperature and pressure over the range of interest. Figure 11(a) shows the strong influence of temperature on density for the temperature range from 30.8 to 41.3 °R at a constant pressure of 30 psia. The minor influence of pressure on LH_2 density for a constant temperature of 36 °R over the pressure range from 15 to 50 psia is also shown in figure 11(b). The uncertainty in density determination was estimated by equation (3).

$$U\rho = (\partial \rho / \partial T * UT) \tag{3}$$

where $U\rho$ is the uncertainty in density, UT is the uncertainty in temperature, and $\partial\rho/\partial T$ is the change in density as a function of temperature. The uncertainty in temperature measurement over the critical LH₂ range for a $1000-\Omega$ ice-point PRT was estimated to be ≤ 0.2 °R. A density uncertainty of ± 0.2 percent, based on the temperature measurement uncertainty of ± 0.2 °R, was estimated for mass flow rate measurements. Pressure measurement error was considered insignificant for LH₂ conditions.

Liquid hydrogen flow rate by orifice-differential pressure measurement error analysis.—Variable reluctance differential pressure transducers have been tested at cryogenic temperatures (ref. 9). The estimated LH_2 mass flow-rate error analysis based on the measurement of differential pressure drop across a flow-control orifice by use of a similar variable reluctance differential pressure transducer is given in table 10(a). An orifice and pressure transducer calibration inaccuracy of ± 0.5 percent of reading for each component was assumed for this study, and the estimated electronics error of the EDU was assumed to be ± 0.4 percent. These values are estimates, and actual LH_2 calibration of the system would have to be performed.

Turbine flow-measurement error analysis.—Table 10(b) lists the estimated flow-measurement error for a commercial turbine flowmeter assumed calibrated at an accuracy of ± 0.5 percent of reading. The assumed EDU measurement inaccuracy is 0.4 percent of range and the estimated density error is ± 0.2 percent.

Target meters.—The feasibility of using a commercially available target flowmeter was investigated. The pressure drop developed across the target of the investigated meter was proportional to the density of the fluid and the volumetric flow rate squared. The meter had a designed full-scale pressure drop of 7 psi that was too high for the COLD—SAT requirements, and the meter would have to be used at its low operational end where its accuracy would be low.

Proposed LH₂ flow measurement system.—Venturi or orifice flow-control devices will be sized to provide the required flow rates. The differential pressure developed across the flow-control devices will be monitored by variable-reluctance differential-pressure transducers. Variable-reluctance transducers have been tested successfully at cryogenic temperatures by a number of facilities and have been selected for liquid helium flow-rate measurements for the superfluid helium on-orbit transfer experiments (SHOOT) (ref. 10). A turbine meter located immediately downstream of the supply tank will be used for measurement redundancy and can be disconnected from the transfer line during line chilldown periods to protect it from overspeeding.

THERMODYNAMIC VENT SYSTEM FLOW-RATE MEASUREMENT REQUIREMENTS

The thermodynamic vent system (TVS) flow rates will be monitored to determine system operation status and to quantify the hydrogen mass expelled. The flow rates will be obtained by the use of flowrestriction devices. The system must be extensively calibrated so that the mass flow will be known as a function of system pressure and temperature. The TVS systems will be designed to provide single-phase hydrogen vapor at their exit. Additional gaseous TVS flow-rate measurements are requested with an inaccuracy less than ± 5.0 percent. The TVS vent plumbing design use tee connections located on the radiator tray where the individual tank vent lines join to a common TVS vent line. The individual tank TVS flowmeters could be located on the radiator tray and must be capable of operating within a temperature range 36 to 360 °R.

Proposed gaseous hydrogen TVS flow-measurement candidate.—The recommended volumetric flowmeter for the TVS systems is the turbine flowmeter. Commercially available turbine meters are designed for both liquid and gas measurements and are operable down to temperatures of 30 °R.

The turbine meter consists of a rotor mounted by bearings inside the meter housing. Fluid flow through the meter imparts a torque to the rotor blades causing the rotor to spin at a rate proportional to volumetric flow rate. Rotation rate is detected by a magnetic or carrier modulated pickup assembly that detects a change in motional inductance of the coil either as a voltage pulse or as a change in carrier frequency. The meters are calibrated and a proportionality constant (K) relating meter output frequency to volume is determined. The deviation of the K factor over the measurement range of the meter from its nominal value is given in the linearity specification. Turbine flowmeters are available with linearity capabilities of ± 1.0 percent of full scale over a normal 10:1 flow range. The calibration of the meters under gaseous hydrogen at the actual operating temperature and pressure conditions will be required. Table 11 lists the various COLD-SAT tank TVS flow rate ranges as a function of temperature and pressure. The required size flowmeters are listed that have operating ranges comparable with the TVS ranges.

TVS flow-measurement density error analysis.—Turbine meters measure volumetric flow rate. In order to calculate mass flow rate, the density must be determined. Temperature and pressure measurements made at the meter inlet will be used to determine density. Mass flow-rate error is a function of the uncertainty of the turbine volumetric flow measurement and of the uncertainty in density determination based on the temperature and pressure measurement errors.

The influence of temperature and pressure on GH_2 density was estimated by finding the best fit equations that mathematically describe the change in hydrogen density as a function of temperature and pressure over the range from 180 to 360 °R and 5 to 20 psia. Figure 12 shows that the hydrogen density is proportional to the pressure-to-temperature ratio (ideal gas law). The uncertainty in density determination was calculated using equation (4). The results of this analysis indicate that a rss density uncertainty of 3.66 percent could occur.

$$U\rho = \left[\left(\frac{\partial \rho}{\partial \Gamma} \times UT \right)^2 + \left(\frac{\partial \rho}{\partial P} \times UP \right)^2 \right]^{0.5}$$
 (4)

where uncertainty in density is $U\rho$, $\partial\rho/\partial T$ is the change in density as a function of temperature, $\partial\rho/\partial P$ is the change in density as a function of pressure, UT is the uncertainty in temperature, and UP is the uncertainty in pressure.

TVS flow-measurement error analysis.—Table 12 lists the error contributions of the various sources involved in the TVS flow measurement. The rss combination is listed and the results show the feasibility of a ± 5 percent mass flow-rate measurement.

Supply and Receiver Tanks Vent Flow-Rate Measurement Requirement

The capability of measuring the vent flow rates of the supply and receiver tanks is required for fluid inventory management and receiver tank chilldown performance. Tank vent flow rates of approximately 50 lbm/hr are estimated.

Proposed tank vent flow-rate measurement system.—A flow-regulation device will be required to control the rate of tank pressure decrease so that valve actuation and tank pressure can be controlled. The tanks will be vented at pressures of approximately 50 psia to a near vacuum condition. This large pressure drop across a sonic flow nozzle will result in a choked flow condition at the throat section of the nozzle. The velocity of the vented gas will reach the sonic value and will not be influenced by changes in downstream pressure (ref. 11). The flow rate is dependent only on the upstream pressure. Temperature of the vented gas will be measured so that density and mass flow rates can be calculated. Commercial suppliers of sonic flow nozzles have designed flow nozzles that develop sonic flow with downstream pressures as high as 80 percent of the nozzle inlet pressure.

Sonic flow-nozzle sizing.—The required diameter of the nozzle can be estimated as follows:

$$M = \rho A V_{s} \tag{5}$$

where M is the mass flow rate, ρ is the density, A is the area of nozzle throat, and V_s is the sonic velocity. Table 13 lists the properties of GH_2 and their influence on the required diameter of a sonic nozzle for a mass flow rate of 50 lbm/hr (0.0139 lbm/sec) at a inlet pressure of 50 psia.

Large receiver tank chilldown vent flow-rate measurement.—The receiver tanks must be chilled down before a LH₂ transfer experiment. The proposed method of tank chilldown is to inject a known mass of LH₂ into the tank. The LH₂ absorbs heat, vaporizes and pressurizes the tank. The tank will be vented when the pressure reaches 30 psia. The desired rate of pressure decrease is 5 psia in 10 sec. This charge-hold-vent process will be repeated until the tank temperature decreases from an initial value of >400 °R to the target temperature of approximately 140 °R. The amount of mass that must be vented from the large receiver tank to lower its pressure from 30 to 25 psia is listed as a function of temperature in table 14. The approximate time required to vent this amount of hydrogen through a 0.151-in.-diameter sonic nozzle is also listed.

Small receiver tank vent flow-rate measurement.—The volume of the small receiver is 13.5 ft³ and that of the large receiver 21 ft³. The volume ratio of the two tanks is 0.643. To acquire the same rate of pressure decrease as the large receiver tank, the small receiver vent line would require a sonic nozzle with a diameter of $(0.643)^{1/2}$ that of the large receiver, or 0.121 in.

Sonic nozzle flow-measurement error analysis.—The inaccuracy in the vent flow-rate measurements is dependant on the calibration error of the nozzle, typically ± 0.75 percent, and on the accuracy of the required absolute pressure and temperature measurements at the entrance to the sonic nozzle. The general formula for mass flow rate M through the nozzle is

$$M = KP/\left(T\right)^{1/2} \tag{6}$$

where P is the absolute pressure (in pounds per square inch) at the nozzle inlet, K is the flow coefficient, and T is the nozzle inlet temperature. The uncertainty in the flow measurement UM was estimated as follows:

$$UM = \left\{ \left[\left(\frac{\partial M}{\partial P} \right) UP \right]^2 + \left[\left(\frac{\partial M}{\partial T} \right) UT \right]^2 + UM_{\text{cal}}^2 \right\}^{0.5}$$
 (7)

where $\partial M/\partial P$ is the change in flow with respect to pressure, UP is the uncertainty in pressure, $\partial M/\partial T$ is the change in flow with respect to temperature, UT is the uncertainty in temperature, and UM_{cal} is the uncertainty in flow calibration. The results of this analysis using the pressure and temperature measurement uncertainties obtained with the proposed PRT temperature sensors and EDU measured 50-psia range pressure transducers are listed in table 15. The results indicate that a < ± 5 percent measurement is feasible.

Acceleration measurement requirement.—The influence of acceleration levels on the fluid dynamic and thermodynamic properties will be determined during the tank chilldown experiments by accelerating the spacecraft at levels from 10 to 100 μg . The other COLD–SAT experiments will be conducted at acceleration levels where natural convection is insignificant ($\leq 10 \ \mu g$). The acceleration magnitude along all three axes will be measured.

Recommended accelerometer system.—A transducer assembly similar to that used on the space shuttle orbiter program is recommended for COLD-SAT. This assembly consists of three accelerometers and the required power conditioning, analog servoelectronics, signal conditioning and electromagnetic interference electronics. The unit has been used on the space shuttle orbiter programs to measure low level accelerations in the micro-g range to 10 milli-g.

The three-axis 0- to 5-V output signals of the accelerometer would be measured at the EDU boxes and converted to their digital equivalent levels by the high-resolution 12-bit analog-to-digital converters. A peak hold circuit will register peak transient acceleration levels. The peak detector will be read and reset at a 1-sec rate.

Mixer Flow-Rate Measurement Requirement

A mixer is required for the active thermodynamic vent system of the supply tank. The mixer pump must be capable of supplying LH₂ at flow rates of 3.0 to 12.7 gal/min. The flow rate of the mixer can be obtained by calibrating the mixer shaft rotation rate with respect to flow rate. The mixer shaft rotation rate can be detected by induction or magnetic type pulse detecting sensors.

A speed sensor compatible with the LH₂ environment will be supplied with the selected mixer pump. The frequency pulse output of the mixer speed sensor will be proportional to the shaft speed and the flow rate. The frequency output of the speed sensor is supplied to an EDU for signal conditioning and fed to a RCTU/CTU of the TT&C system. Figure 13 shows a block diagram of the rotational speed measurement system. This design is similar to a rotational speed measurement design used in the shuttle/ Centaur design.

Valve Status Indicators

The cryogenic and standard temperature rated valves proposed for the COLD-SAT plumbing system are manufactured with valve position indicating switches. The status of the switches (open/closed) will be monitored by the remote command telemetry units to verify operational status. Only 1 bit is required for digitizing this information.

Figure 14 shows the typical schematic for discrete measurements. This design is similar to the Shuttle/Centaur design. The proposed design uses manganin wiring for the cryogenic valve switches to minimize heat conduction.

Power Measurements

The voltage and current supplied to the COLD-SAT experiment system heaters and mixers will be measured to determine the power dissipated by the components. The measurements will be made at the relay sequencer units of the TT&C system.

COLD-SAT EXPERIMENT SUBSYSTEM MODULE INSTRUMENTATION REQUIREMENTS AND ALLOCATIONS

The following sections present brief descriptions of the COLD-SAT experiments and lists the number of transducers allocated to each COLD-SAT module to meet the experiment requirements.

Supply Module Instrumentation Requirements

The instrumentation allocated for the supply module was dictated by the requirements of the following experiments.

Experiment 1: tank pressure control experiment.—The pressurization rate of a cryogenic storage tank depends on the rate and distribution of heat energy into the tank. In space the lack of gravity-induced fluid convection currents can result in nonuniform heating of the cryogen and the formation of temperature stratified layers of fluid. The ullage pressure will be determined by the warmest layer of liquid and the tank pressurization rate will be high.

The tank pressure-control experiments will investigate the rate of tank pressurization as a function of heat flux, tank fill, and acceleration levels. Methods to reduce thermal stratification by means of active and passive thermodynamic vent systems (TVS) will be evaluated.

Key measurements.—Quantification of the heat energy entering the tank will be determined by measuring the temperatures and temperature gradients existing throughout the supply tank module. The internal tank fluid state (liquid or vapor) and temperatures will be monitored using specially designed probes for low-gravity applications. Absolute tank pressure and differential pressure drops across flow-control devices, acceleration level, mixer motor speed, valve status, and heater power levels will be monitored.

Table 16 list the type, number, and location of the instrumentation proposed for the supply tank module. A measurement list detailing the description, range, accuracy, and sample rate for each measurement is given in the appendix. The supply tank module contains the LH₂ supply tank and gas pressurization system. The accelerometer is listed under this module for convenience only. The accelerometer will be located in the midelectronic bay.

Receiver Modules Instrumentation Requirements

The instrumentation proposed for the large receiver and the small receiver modules was dictated by the requirements of the following experiments.

Experiment 2: No-vent fill of cryogenic tanks in low gravity.—The objectives of the no-vent fill experiments were to characterize the influence of the thermodynamic state of the liquid, the liquid flow rate, and the liquid-injection technique on the maximum amount of tank liquid fill that can be obtained.

Experiment 3: Cryogenic tank chilldown in low gravity.—The first step in the LH₂ transfer process is to chill down the receiver tank to a temperature at which a continual no-vent-fill fluid transfer can be obtained. Tank chilldown experiments will consist of injecting known charge magnitudes of LH₂ into the receiver tank. The LH₂ will vaporize and absorb heat from the receiver tank and thermally bonded structure. After a sufficient hold time the tank will be vented and the remaining fluid will again absorb heat from the structure. This process of charge-hold-vent will be repeated until a target temperature is reached, at which point a no-vent fill transfer could be performed. The influence of LH₂ charge flow rate, acceleration, and spray method will be evaluated to determine the optimum method of tank chilldown.

Experiment 4: Fill of liquid acquisition devices in low gravity.—The capability of filling a warm LAD and refill a cold LAD will be performed on the large receiver tank. Tables 17 and 18 list the type, number, and location of the instrumentation selected for the large and small receiver modules. A measure-

ment list containing the description, range, accuracy, and sample rate for each large and small receiver module measurement is located in the appendix.

Instrumentation Redundancy

Large quantities of temperature, pressure, and liquid-vapor point sensors have been allocated to the COLD-SAT tank modules. The loss of any one specific sensor would not significantly jeopardize the experimental measurement data acquisition or accuracy. Three absolute pressure transducers are assigned to each tank for voting purposes. High-accuracy LH₂ flow measurements will be determined by differential pressure measurements across the flow-control elements. A turbine flowmeter is also assigned for secondary total flow measurement. The thermodynamic vent system and vent lines will be instrumented, and flow rates will be determined from calibration data as a function of temperature and pressure. Turbine flowmeters will also provide TVS flow-rate measurements.

The instrumentation for each module is distributed equally among the three command telemetry units and the three EDU's so that if a signal conditioning board failure occurs, sensor loss would be distributed among the three modules so as to minimize the loss of experimental data.

COLD-SAT EXPERIMENT INSTRUMENTATION AND ELECTRICAL HARNESS DESIGN

Electrical wiring is required to supply the excitation power to the sensors and to couple the sensor output signal to the DAS's. Most of the instrumentation will be thermally bonded to the cryogenic tanks and a large heat input to the tanks would occur if copper wire is used. To minimize heat input, manganin wire was chosen for instrumentation wiring. Phosphor bronze wiring was chosen for valves and heater applications. The influence of temperature on the thermal conductivity of copper, manganin, and phosphor bronze is shown in figure 15.

Connectors and receptacles used for cryogenic tank wire feedthroughs must be capable of operation at LH₂ temperatures. The candidate connector assemblies will be similar to the type used on D-1A Centaur. Bulkhead connector assemblies exposed to temperatures below -55 °F will be series 40M38294. Connectors exposed to temperatures above -55 °F will be series MIL-C-38999. Detailed descriptions of the designed experiment system instrumentation and electrical wire harness is included in the following sections.

Wire Material Selection

Copper wire is an excellent electrical conductor and also a very good thermal energy conductor. However, the large number of instrumentation wires thermally bonded to the COLD-SAT supply tank would result in excessive conducted heat input and boiloff of LH₂ during the mission. Copper wire, therefore, is not practical, and a material with a lower thermal conductivity is desired. Manganin, a copper alloy consisting of 87 percent copper, 11 percent manganese, and the remainder nickel, has a much lower thermal conductivity than copper. Manganin is commonly used for cryogenic temperature measurements to prevent thermal conduction temperature measurement errors. The resistivity of manganin is considerably greater than copper and four-wire measurements where two wires supply power to the sensor and two wires sense the potential across the sensor must be used for accurate results. The heat conducted down the additional manganin wiring is still far lower than that conducted down two wire copper systems.

Manganin wire of 24 AWG size has been selected for the cryogenic instrumentation wiring harnesses. The current requirements of the selected instrumentation is generally 10 mA or less, and the 24 AWG gauge size should be more than adequate. This gauge size should also provide the strength and maintenance properties desirable for interconnection harnessing.

Valves and heaters will require larger diameter wire because of their greater current carrying requirements (1 to 5 A). Manganin wire of 12 to 10 gauge would be required to provide this current carrying capacity. This size wire would present a large weight increase.

Phosphor bronze is a copper wire alloy commonly used in cryogenic applications. This wire is beneficial for cryogenic applications requiring current carrying capacities unpractical for manganin. 18—AWG phosphor bronze wire was selected to represent the typical wire material for valve and heater leads that will be thermally bonded to the cryogenic tanks.

Wire Construction Model

The wire construction was modeled after MIL-W-81381, which specifies a fluorocarbon and polyimide insulation rated for +200 °C. The estimated weight of 24 AWG manganin and 18 AWG phosphor bronze wire based on the specification data is listed in table 19. Table 20 lists the estimated diameters of the various wire bundles.

Instrumentation and Power Wire Requirements

Table 21 is a summary of the required type and number of wire for each COLD-SAT tank module. Tables 22 to 24 list the number and termination locations of the supply, large receiver, and small receiver module wiring.

Signal Conditioner Assignments

The instrumentation and electrical actuated components of the modules were divided between the various signal conditioning units and sequencer boxes so as to minimize the loss of measurement data in case of a harness or electronic system failure. Tables 25 and 26 list the wire allocations.

Cryogenic Tank Wire Connectors and Feedthroughs

The internal tank instrumentation wiring will penetrate the pressure vessels through hermetically sealed electrical wire feedthroughs. The connector style selected is similar to the type used on D-1A Centaur. The desired modification is to have the feedthrough assembly supplied in a weld-mount style. Six feedthrough assemblies are required for the supply tank: four for the large receiver, and three are required for the small receiver.

Multiplexer Wire Reduction Techniques

A large number of temperature and liquid/vapor sensors are installed inside the cryogenic tanks. The sensors require two to four wires each. The large number of wires penetrating the cryogenic tanks increases the conducted heat leak and require large numbers of cryogenic tank feedthroughs.

Methods of using four-channel differential joint Army Navy (JAN)-qualified analog multiplexers were developed that use common excitation lines to reduce the number of wires penetrating the tanks. Figure 16(a) is a schematic of a multiplexed four-wire temperature measurement method in which the required number of wires is reduced from 64 to 40 wires for the 16 sensors.

Large numbers of liquid-level point detectors are also required to determine the fluid-phase distribution throughout the COLD-SAT tanks. To reduce the number of wires penetrating the tanks, the sen-

sors will be connected in parallel to common excitation lines inside the cryogenic tanks. Figure 16(b) shows how the multiplexers could be used to select and excite the desired sensor to be measured. The number of wires penetrating the tanks by this method is 1/4 the number required for individually wired detectors.

Wire Harness Construction

The COLD-SAT experiment subsystem will be assembled in separate tank modules. The instrumentation will be mounted on the LAD instrumentation rakes and checked for proper operation before mounting the LAD in the tank. The internal tank instrumentation wiring will be connected to the cryogenic plug, and electrical contact to the external harness will be made through the cryogenic receptacle. The instrumentation and power wiring harnesses will be separated and routed from the tank structure to the radiator tray in bundles of 40 to 55 wires. The radiator tray is located on the cold side of the spacecraft and is designed to serve as a heat sink to remove heat energy from the wiring prior to its arrival to the tank. Connector feedthroughs will exist at panels located at module interfaces. The temperature of the radiator tray may be below –100 °F and cryogenic environment resisting connectors as specified in 40M38294 will be required.

The wire harnesses will be routed from the tray to bulkhead connectors located on panels located on the warm side of the spacecraft. This design is to prevent undesirable chilldown of the electronic boxes. From the warm-side panels, the harnesses terminate at their required electronic or power box. Spacecraft warm-side electrical connectors will be selected from Mil-C-38999. Figure 17 shows the overall wire schematic for the COLD-SAT experiment subsystem instrumentation.

Electromagnetic Interference Concerns

The instrumentation and power wiring should be separated and routed away from noise sources. The wire should be twisted to minimize inductive noise coupling. The low-level signal wiring should be harnessed in shielded bundles until it is inside the multilayer insulation can.

Heater Requirements

Heaters are required to provide the desired experiment heat flux inputs for the tank pressurization experiments. Heaters are also required to warm the receiver tanks to the required initial temperatures before childown experiments. The hydrogen vaporizers will use heaters to warm and maintain their system at their designed operating point. Panel and vent line heaters will be required to maintain their structure at desired operating conditions.

Thin, flexible heaters consisting of etched metal foil resistive elements laminated between layers of Kapton insulation are commercially available. Goddard Space Flight Center has qualified similar heaters for aerospace applications from -65 to 200 °C.

The listed lower temperature range of operation of commercially available Kapton-insulated heaters is -200 °C. Discussions with manufacturing representatives indicate that the heaters should be capable of operation at LH₂ temperatures as long as appropriate installation adhesives and techniques are used. Testing would be required to qualify an adhesive for this application.

CONCLUSIONS

An instrumentation and wire harness for the COLD-SAT experiment subsystem was designed. The transducers, signal conditioning systems, and wire harness components of the design were recommended, when at all possible, based on past LH₂ and space-flight histories. Electrical current excitation levels and data acquisition ranging were designed to meet the temperature measurement requirements using space flight qualified platinum resistance thermometers with minimal changes to existing 8-bit space flight qualified data acquisition systems. Cone shaped LH₂ temperature and liquid/vapor sensors were recommended to minimize false measurements in low-gravity conditions. The need for a 12-bit data acquisition system to provide improve resolution and accuracy for critical measurements was addressed, and an EDU design was presented. The 12-bit system was identical to the 8-bit system but with only one multiplexer for each circuit card and the 12-bit analog-to-digital converter. Error analyses were performed on all instrument candidates, and the influence of physical parameters such as temperature and pressure on the overall measurement accuracies was investigated. The results indicated that most of the experimenters' measurement requirements were feasible; however, two-phase flow detection, leak detection systems, and mass gauging are areas that need development.

Heat conduction to the LH₂ tanks was a chief concern. Wire materials, wire multiplexing techniques, and cryogenic operable pressure transducers were selected and developed to minimize this problem.

REFERENCES

- 1. Zuch, E.L.: Data Acquisition and Conversion Handbook: A Technical Guide to A/D D/A Converters and their Applications. Datel Intersil, Ma, 1979.
- 2. Abernethy, R.B.; Benedict, R.P.; and Dowdell, R.B.: ASME Measurement Uncertainty. J. Fluids Eng., vol. 107, June 1985, pp. 161–164.
- 3. Weast, R.C.; Astel, M.J.; and Beyer, W.H., eds.: CRC Handbook of Chemistry and Physics. 68th ed., CRC Press, Inc. Boca Raton, FL, 1987, p. E-9.
- 4. Touloukian, Y.S., Powell, R. W.; Ho, C.Y.; and Klemans, P.G.: Thermophysical Properties of Matter. Vol. 1: Thermal Conductivity, Metallic Elements and Alloys. IFI/Plenum, Washington, DC, 1970.
- 5. McCarty, R.D.: Hydrogen Technological Survey: Thermophysical Properties. NASA SP-3089, 1975.
- 6. Liptak, B.G.; and Venczel, K.: Instrument Engineers Handbook, Rev. ed., Chilton Book Co., Radnor, PA, 1982.
- 7. Berns, J.A.; Yeh, F.C.; and Nussel, R.C.: Photographic Investigation of Liquid Behavior on Temperature and Liquid/Vapor Sensors Used in Low-Gravity Environment. NASA TM X-1438, 1967.
- 8. Baird, R.S.: Flowmeter Evaluation For On-Orbit Operations. NASA TM-100465, 1988.
- Kashani, A.: Shoot Flowmeter and Pressure Transducers. Cryogenics, Vol. 30, Mar. 1990, pp. 286–291.
- 10. DiPirro, M.J.; and Kittel, P.: The Superfluid Helium On-Orbit Transfer (SHOOT) Flight Experiment. Advances In Cryogenic Engineering, Vol. 31, Plenum Press, New York. 1988.
- 11. John, J.E.A.; and Haberman, W.L.: Introduction to Fluid Mechanics. Second ed., Prentice-Hall, Inc., Englewood Cliffs, NJ, 1980.

APPENDIX—COLD-SAT EXPERIMENT SUBSYSTEM MEASUREMENT LISTS:

Measurement lists containing the descriptions, ranges, inaccuracies and sampling rates for the COLD-SAT experiment measurements were prepared. The measurements are referenced to the primary COLD-SAT experiment requiring the measurement. Measurement numbers are assigned according to the definitions and nomenclature discussed below.

COLD-SAT EXPERIMENT SUBSYSTEM MEASUREMENT LIST NOMENCLATURE AND DEFINITIONS:

MEASUREMENT LOCATIONS MEASUREMENT TYPES

 $S = Supply \ Module \\ Lr = Large \ Receiver \\ Sr = Small \ Receiver \\ E = Voltage \\ I = Current \\ C = Capacitance \\ R = Rate \ (RPM) \\ T = Temperature$

W = Power

PRIMARY EXPERIMENTS: 1 = PRESSURE CONTROL 2 = NO-VENT FILL

3 = CHILL DOWN 4 = LAD

8 = PRESSURIZATION

RANGE: The listed experiment range is based on the expected state of the fluid over the course of

the experiment.

INACCURACY: The listed inaccuracy is based on the Root Sum Square contributions of sensor

inaccuracy and estimated data acquisition error over the measurement range. This result is an estimate and all sensors should be calibrated and evaluated under

actual operating conditions.

SAMPLE RATE: The sample rate represents the frequency at which the measurement will be

made. The actual sampling time will be considerably less.

TEMPERATURE MEASUREMENT IDENTIFICATION CODE: Temperature sensors were classified according to their range of operation. Type A sensors cover the temperature range from 29 °R to 46 °R. Type B sensors cover the temperature range from 113 °R to 540 °R. Sensors classified as type AB indicate dual ranging of the sensor which extends the range of coverage from 29 °R to 540 °R but at reduced accuracy.

To facilitate locating temperature sensors in the various regions of the tank modules, the sensors were numbered by the following method. Temperature measurement numbers ranging from 1 to 99 indicate measurements located inside the tank pressure vessel. Temperature measurement numbers ranging from 101 to 199 indicate measurements located on the outside pressure vessel wall. Temperature measurement numbers ranging from 201 to 299 indicate measurements located between the multilayer insulation (MLI) and the pressure vessel. Temperature measurement numbers ranging from 301 to 399 indicate measurements located on or outside the MLI structure.

CRYOGENIC TANK INTERNAL FLUID TEMPERATURE MEASUREMENTS: The fluid temperature distribution existing throughout the cryogenic tanks will be measured using Platinum Resistance Thermometers (PRTs). The fluid temperature sensors will be mounted in cone shaped probes similar to those used on Centaur. These probes are designed to wick away liquid films from the sensor tip, minimizing false ullage temperature readings.

The temperature probes in the supply tank are mechanically secured to (7) seven instrumentation rakes. Horizontal rakes are located at 70% and 30% tank volume levels. These rakes will monitor the fluid temperature distribution existing from the tank wall into the fluid bulk. A vertical rake positioned at the 50% volume level will monitor the liquid/vapor interface temperature during a low gravity settling period. The fluid temperature between 90% to 98% volume levels will be measured by temperature probes mounted on vertical rake #5. Four LAD mounted rakes will contain sensors to determine the temperature distribution throughout the 4% to 90% tank volume region.

A minimum LH2 temperature of 30.8° should occur downstream of the Joule Thomson expansion device. The maximum Tank pressure should not exceed 50 PSIA. The H2 saturation temperature for this pressure is 45.5 °R. Type A PRTs are excited and signal conditioned to maximize their sensitivity over this temperature range.

QUANTIFYING THE HEAT ENERGY INPUT INTO THE CRYOGENIC TANKS: The heat energy conducted into the cryogenic tanks will be determined by monitoring the temperature and temperature gradients existing across the module components that are thermally bonded to the tanks. These components consist of the plumbing penetrations, struts, and wire harnesses. Temperature maps of the tank surface, inner MLI surface, and outer MLI surfaces will be obtained to determine the thermal radiative environment of the module.

TANK PANEL LOCATED TEMPERATURE MEASUREMENTS: Aluminum panels are mounted on the surface of the supply and receiver tanks. The panels support plumbing and instrumentation components. The panels are thermally bonded to the tanks and the temperature of the panels and their components will be in the Type A temperature measurement range.

INNER HONEYCOMB WALL TEMPERATURE MEASUREMENTS: The temperature range listed for the inner MLI-Honeycomb structure were determined from the thermal analysis. The range covers the estimated worse case deviations in temperature.

OUTER MLI TEMPERATURE MEASUREMENTS: The temperature range listed for the outer MLI structure was determined from the thermal analysis. The range covers the worse case deviations in temperature.

SUPPLY AND RECEIVER MODULES INSTRUMENTATION LOCATION SCHEMATICS: Figures 18(a) and (b) show the location of the supply tank rake mounted temperature and liquid-vapor sensors. Figure 19 shows the instrumentation rake design and external sensor locations for the large receiver. Figure 20 shows the instrumentation rake design and the external sensor locations for the small receiver.

COLD-SAT EXPERIMENT SUBSYSTEM MEASUREMENT LIST

SUPPLY TANK INTERNAL TEMPERATURE MEASUREMENTS

 MEAS	NUM	DESCRIPTION	RANGE Deg (R)	INACCURACY ± Deg (R)	SAMPLE FREQ. (Hz)	PRI. EXP.
STa	1	Horizontal Rake 70% Volume Level	29 - 46	< 0.20	0.1	1
STa	2	Horizontal Rake 70% Volume Level	29 - 46	< 0.20	0.1	1
STa	3	Horizontal Rake 70% Volume Level	29 - 46	< 0.20	0.1	1
STa	4	Horizontal Rake 70% Volume Level	29 - 46	< 0.20	0.1	1
STa	5	Horizontal Rake 70% Volume Level	29 - 46	< 0.20	0.1	1
STa	6	Horizontal Rake 70% Volume Level	29 - 46	< 0.20	0.1	1
STa	8	Horizontal Rake 30% Volume Level	29 - 46	< 0.20	0.1	I
STa	9	Horizontal Rake 30% Volume Level	29 - 46	< 0.20	0.1	1
STa	10	Horizontal Rake 30% Volume Level	29 - 46	< 0.20	0.1	1
STa	11	Horizontal Rake 30% Volume Level	29 - 46	< 0.20	0.1	1
STa	12	Horizontal Rake 30% Volume Level	29 - 46	< 0.20	0.1	1
STa	13	Vertical Rake 50% Volume Level	29 - 46	< 0.20	0.1	1
STa	14	Vertical Rake 50% Volume Level	29 - 46	< 0.20	0.1	1
STa	15	Vertical Rake 50% Volume Level	29 - 46	< 0.20	0.1	1
STa	16	Vertical Rake 50% Volume Level	29 - 46	< 0.20	0.1	1
STa	17	Vertical Rake 50% Volume Level	29 - 46	< 0.20	0.1	1
STa	18	Vertical Rake 50% Volume Level	29 - 46	< 0.20	0.1	1
STab	19	Rake #1 90% Volume Level	29 - 46	< 0.20	0.1	1
STa	20	Rake #1 80% Volume Level	29 - 46	< 0.20	0.1	1

MEAS	NUM	DESCRIPTION	RANGE Deg (R)	INACCURACY ± Deg (R)	SAMPLE FREQ. (Hz)	PRI. EXP.
STa	21	Rake #1 72% Volume Level	29 - 46	< 0.20	0.1	1
STa	22	Rake #1 56% Volume Level	29 - 46	< 0.20	0.1	l
STa	23	Rake #1 48% Volume Level	29 - 46	< 0.20	0.1	1
STa	24	Rake #1 32% Volume Level	29 - 46	< 0.20	0.1	1
STa	26	Rake #2 88% Volume Level	29 - 46	< 0.20	0.1	I
STa	27	Rake #2 78% Volume Level	29 - 46	< 0.20	0.1	1
STa	28	Rake #2 68% Volume Level	29 - 46	< 0.20	0.1	1
STaB	29	Rake #2 54% Volume Level	29 - 46	< 0.20	0.1	1
STa	30	Rake #2 44% Volume Level	29 - 46	< 0.20	0.1	1
STa	31	Rake #2 28% Volume Level	29 - 46	< 0.20	0.1	I
STa	32	Rake #2 12% Volume Level	29 - 46	< 0.20	0.1	I
STa	33	Rake #3 86% Volume Level	29 - 46	< 0.20	0.1	1
STa	. 34	Rake #3 76% Volume Level	29 - 46	< 0.20	0.1	1
STab	35	Rake #3 64% Volume Level	29 - 46	< 0.20	0.1	1
STa	36	Rake #3 52% Volume Level	29 - 46	< 0.20	0.1	1
STa	37	Rake #3 40% Volume Level	29 - 46	< 0.20	0.1	1
STa	38	Rake #3 24% Volume Level	29 - 46	< 0.20	0.1	1
STa	39	Rake #3 8% Volume Level	29 - 46	< 0.20	0.1	1
STab G	40	Rake #4 84% Volume Level	29 - 46	< 0.20	0.1	1
STab G	41	Rake #4 74% Volume Level	29 - 46	< 0.20	0.1	1

MEAS	NUM	DESCRIPTION	RANGE Deg (R)	INACCURACY ± Deg (R)	SAMPLE FREQ. (Hz)	PRI. EXP.
STa G	42	Rake #4 60% Volume Level	29 - 46	< 0.20	0.1	1
STab G	43	Rake #4 50% Volume Level	29 - 46	< 0.20	0.1	1
STa G	44	Rake #4 36% Volume Level	29 - 46	< 0.20	0.1	1
STa G	45	Rake #4 20% Volume Level	29 - 46	< 0.20	0.1	1
STa G	46	Rake #4 4% Volume Level	29 - 46	< 0.20	0.1	1
STa G	47	Rake #5 98% Volume Level	29 - 46	< 0.20	0.1	i
STa G	48	Rake #5 95% Volume Level	29 - 46	< 0.20	0.1	1
STab G	49	Rake #5 92% Volume Level	29 - 46	< 0.20	0.1	I
STab	50	Press diffuser	29 - 46	< 0.20	0.1	1
STa	51	Forward LAD box	29 - 46	< 0.20	0.1	1
STa	52	AFT LAD box	29 - 46	< 0.20	0.1	1
STa	53	LAD channel #I 1/3 length	29 - 46	< 0.20	0.1	1
STa	54	LAD channel #1 2/3 length	29 - 46	< 0.20	0.1	1
STa	55	LAD channel #2 1/3 length	29 - 46	< 0.20	0.1	1
STa	56	LAD channel #2 2/3 length	29 - 46	< 0.20	0.1	1
STa	57	LAD channel #3 1/3 length	29 - 46	< 0.20	0.1	1
Sta	58	LAD channel #3 2/3 length	29 - 46	< 0.20	0.1	1
STa	59	LAD channel #4 1/3 length	29 - 46	< 0.20	0.1	1
STa	60	LAD channel #4 2/3 length	29 - 46	< 0.20	0.1	1
Sta	61	1/4 inch from TVS line, Channel #1	29 - 46	< 0.20	0.1	1
STa	62	1/4 inch from TVS line, Channel #2	29 - 46	< 0.20	0.1	1

MEAS	NUM	DESCRIPTION	RANGE Deg (R)	INACCURACY ± Deg (R)	SAMPLE FREQ. (Hz)	PRI. EXP.
STa	63	1/4 inch from TVS line, Channel #3	29 - 46	< 0.20	0.1	1
Sta	64	1/4 inch from TVS line, Channel #4	29 - 46	< 0.20	0.1	1
SUPPLY	TANK OU	JTER WALL TEMPERATU	IRE MEASUR	EMENTS		
STa	101	LH2 transfer line temperature at supply tank wall	29 - 46	< 0.20	0.1	I
STa	102	fill drain line temperature at supply tank	29 - 46	< 0.20	0.1	1
Sta	103	passive TVS line temperature at outer supply tank wall	29 - 46	< 0.20	0.1	1
STab	104	gas pressurization line temp at supply tank wall	29 - 46	< 0.20	0.1	I
STa	105	vent line temp- erature at supply tank outer wall	29 - 46	< 0.20	0.1	1
Sta	106	top dome warm side 29 - temperature	46	< 0.20	0.1	1
STa	107	supply tank upper barrel cold side temperature	29 - 46	< 0.20	0.1	1
STa	108	AFT dome cold side 29 - temperature	- 46	< 0.20	0.1	1
Sta	109	supply tank lower barrel warm side temperature	29 - 46	< 0.20	0.1	1
STa	110	top dome cold side temperature	29 - 46	< 0.20	0.1	1
STa	111	AFT dome warm side temperature	29 - 46	< 0.20	0.1	1
Sta	112	forward barrel warm 29 side temperature	- 46	< 0.20	0.1	1
STa	113	aft barrel cold side temperature	29 - 46	< 0.20	0.1	1
Sta	114	central barrel +y direction	29 - 46	< 0.20	0.1	1
STa	115	central barrel -y direction	29 - 46	< 0.20	0.1	1

MEAS	NUM	DESCRIPTION	RANGE Deg (R)	INACCURACY ± Deg (R)	SAMPLE FREQ. (Hz)	PRI. EXP.
STa	116	forward barrel warm side strut near tank wall	29 - 46	< 0.20	0.1	1
STa	117	aft barrel cold side strut near tank wall	29 - 46	< 0.20	0.1	1
STa	118	harness at PV warm side	29 - 46	< 0.20	0.1	1
STa	119	harness at PV cold side	29 - 46	< 0.20	0.1	1
SUPPLY	TANK M	ODULE PANEL LOCATED	TEMPERAT	JRE MEASUREMEN	πs	
STa	201	LH2 transfer line temperature at Flow meter panel J	29 - 46	< 0.20	0.1	I
STa	202	LH2 transfer line temperature at Ventri flow control entrance, Panel I	29 - 46	< 0.20	0.1	1
STa	203	LH2 transfer line temp. at Orifice flow control bank exit. Panel I	29 - 46	< 0.20	0.1	1
Sta	204	LH2 transfer line temp. at panel f entrance. This line section feeds the active TVS mixer pumps	29 - 46	< 0.20	0.1	1
STa	205	LH2 transfer line at Panel F exit. Active TVS mixer pumps exit temp.	29 - 46	< 0.20	0.1	I
STa	206	passive TVS line Temp. at Joule Thompson dev. Panel H	29 - 46	< 0.20	0.1	1
Sta	207	TVs line temp. downstream of JT devices. Panel H	29 - 46	< 0.20	0.1	I
STa	208	Joule Thompson dev. heater temp Panel H	29 - 46	< 0.20	0.1	1
STab	209	gas pressurization line temperature at Panel J	29 - 46	< 0.20	0.1	1
STa	210	ground fill line temperature at Panel E exit	29 - 46	< 0.20	0.1	1

MEAS	NUM	DESCRIPTION	RANGE Deg (R)	INACCURACY ± Deg (R)	SAMPLE FREQ. (Hz)	PRI. EXP.
STa	211	vent line temp. at Panel G	29 - 46	< 0.20	0.1	1
INNER I	HONEYCO	MB TEMPERATURE MEA	SUREMENTS			
STb	212	forward dome warm side temp. of inner honeycome structure	117-280	< 1.5	0.1	1
STb	213	forward barrrel warm honeycomb temp	117-280	< 1.5	0.1	1
STb	214	Aft Dome inner honeybomb temp. cold side	117-280	< 1.5	0.1	1
STb	215	Aft barrel inner honeycomb warm side temperature	117-280	< 1.5	0.1	1
ЅТЪ	216	forward dome cold si honeycomb temp.	117-280	< 1.5	0.1	1
ЅТъ	217	AFT dome warm side honeycomb temp.	117-280	< 1.5	0.1	1
ЅТЪ	218	forward barrel cold side honeycomb temp	117-280	< 1.5	0.1	1
STb	219	Aft barrel cold side honeycomb temp	117-280	< 1.5	0.1	1
STb	220	central barrel +y direction honeycomb temperature	117-280	< 1.5	0.1	1
STb	221	central barrel -y direction honeycomb temperature	117-280	< 1.5	0.1	1
STb	222	Forward barrel warm side strut near honeycomb	117-280	< 1.5	0.1	1
SТЪ	223	Aft barrel cold side strut near honeycomb	117-280	< 1.5	0.1	I
STb	224	harness at honeycomb warm side	117-280	< 1.5	0.1	1
ЅТЪ	225	harness at honeycomb cold side	117-280	< 1.5	0.1	1
HYDRO	GEN VAP	ORIZER TEMPERATURE	MEASUREME	INTS		
STa	301	LH2 transfer line temp. at Vaporizer line inlet	36 - 41	< 0.2	0.1	1

MEAS	NUM	DESCRIPTION	RANGE Deg (R)	INACCURACY ± Deg (R)	SAMPLE FREQ. (Hz)	PRI. EXP.
STab	302	vaporizer B coil inlet temperature	36-530	< 1.5	0.1	1
STab	303	vaporizer A coil inlet temperature	36-530	< 1.5	0.1	I
Stab	304	vaporizer B coil exit temperature	36-530	< 1.5	0.1	1
STab	305	vaporizer A coil exit temperature	36-530	< 1.5	0.1	1
STb	306	vaporizer B bottle cold side temp.	485-525	< 1.5	0.1	1
STb	307	vaporizer A bottle cold side temp.	485-525	< 1.5	0.1	1
STb	308	vaporizer B bottle warm side temp.	485-525	< 1.5	0.1	1
ЅТЪ	309	vaporizer A bottle warm side temp.	485-525	< 1.5	0.1	1
Stab	310	vaporizer coil exit temp. at vent line	38-525	< 1.5	0.1	1
STb	311	helium bottle temperature	485-525	< 1.5	0.1	1
STa	312	helium bottle temperature	485-525	< 1.5	0.1	1
STb	313	helium bottle temperature	485-525	< 1.5	0.1	1
STb	314	helium bottle temperature	485-525	< 1.5	0.1	1
STb	315	Gas pressurant flow meter inlet	485-525	< 1.5	0.1	1
OUTER	MLI TEM	PERATURE MEASUREMI	ENTS			
STb	316	Forward dome outer MLI temperature on warm side	360-525	< 1.5	0.1	1
STb	317	Upper barrel outer MLI temperature on cold side	235-525	< 1.5	0.1	1
STb	318	AFT dome outer MLI temperature on cold side	235-525	< 1.5	0.1	1
SТъ	319	lower barrrel outer MLI temperature on warm side	360-525	< 1.5	0.1	1

MEAS	NUM	DESCRIPTION	RANGE Deg (R)	INACCURACY ± Deg (R)	SAMPLE FREQ. (Hz)	PRI. EXP.				
STb	320	Forward dome outer MLI temperature on cold side	235-525	<1.5	0.1	1				
STb	321	AFT dome outer MLI temperature on warm side	360-525	< 1.5	0.1	ī				
STb	322	upper barrel warm side outer MLI temperature	360-525	< 1.5	0.1	1				
STb	323	lower barrel outer MLI temperature on cold side	235-525	< 1.5	0.1	1				
STb	324	central barrel +y direction outer MLI temperature	235-525	< 1.5	0.1	1				
STb	325	central barrel -y direction outer MLI temperature	235-525	< 1.5	0.1	I				
VENT A	VENT AND GASEOUS FLOW METER TEMPERATURES									
STb	326	ground/ascent vent line temperature	235-525	< 1.5	0.1	1				
STb	327	temperature at vent relief line	235-525	< 1.5	0.1	1				
STb	328	Supply tank active TVS flow meter temp	235-525	< 1.5	0.1	1				
STb	329	Supply tank passive TVS flow meter temp	235-525	< 1.5	0.1	1				
STb	331	gas pressurant flow meter inlet temp.	485-525	< 1.5	0.1	1				
ЅТЪ	332	temperature at balanced TVS vent. (flow)	235-525	< 1.5	0.1	1				
SUPPL	Y TANK L	IQUID VAPOR POINT ME	ASUREMENT	S						
MEAS	NUM	DESCRIPTION	RESPONSE Deg (R)	:	SAMPLE FREQ. (Hz)	PRI. EXP.				
SL	1	Rake #5 98% Volume	LIQ/VAP		0.1	I				
SL	2	Rake #5 96% Volume	LIQ/VAP		0.1	1				
SL	3	Rake #5 95% Volume	LIQ/VAP		0.1	1				
SIL	4	Rake #5 94% Volume	LIQ/VAP		0.1	1				
SL	5	Rake #5 92% Volume	LIQ/VAP		0.1	1				
SL	6	Rake #1 90% Volume	LIQ/VAP		0.1	1				

MEAS	NUM	DESCRIPTION	RESPONSE Deg (R)	SAMPLE FREQ. (Hz)	PRI. EXP.
SL	7	Rake #2 88% Volume	LIQ/VAP	0.1	1
SL	8	Rake #3 86% Volume	LIQ/VAP	0.1	1
SL	9	Rake #4 84% Volume	LIQ/VAP	0.1	1
SL	10	Rake #1 80% Volume	LIQ/VAP	0.1	1
SL	11	Rake #2 78% Volume	LIQ/VAP	0.1	1
SL	12	Rake #3 76% Volume	LIQ/VAP	0.1	1
SL	13	Rake #4 74% Volume	LIQ/VAP	0.1	1
SL	14	Rake #1 72% Volume	LIQ/VAP	0.1	1
SL	15	Rake #2 68% Volume	LIQ/VAP	0.1	1
SL	16	Rake #3 64% Volume	LIQ/VAP	0.1	1
SL	17	Rake #4 60% Volume	LIQ/VAP	0.1	1.
SL	18	Rake #1 56% Volume	LIQ/VAP	0.1	i
SL	19	Rake #2 54% Volume	LIQ/VAP	0.1	1
SL	20	Rake #3 52% Volume	LIQ/VAP	0.1	1
SL	21	Rake #4 50% Volume	LIQ/VAP	0.1	1
SL	22	Rake #1 48% Volume	LIQ/VAP	0.1	1
SL	23	Rake #2 44% Volume	LIQ/VAP	0.1	1
SL	24	Rake #3 40% Volume	LIQ/VAP	0.1	1
SL	25	Rake #4 36% Volume	LIQ/VAP	0.1	I
SL	26	Rake #1 32% Volume	LIQ/VAP	0.1	1
SL	27	Rake #2 28% Volume	LIQ/VAP	0.1	1
SL	28	Rake #3 24% Volume	LIQ/VAP	0.1	1
SL	29	Rake #4 20% Volume	LIQ/VAP	0.1	1
SL	30	Rake #1 16% Volume	LIQ/VAP	0.1	1
SL	31	Rake #2 12% Volume	LIQ/VAP	0.1	1
SL	32	Rake #3 8% Volume	LIQ/VAP	0.1	1
SL	33	Rake #4 4% Volume	LIQ/VAP	0.1	1
SL	34	Vent line entrance	LIQ/VAP	0.1	1
SL	35	Transfer line entrance	LIQ/VAP	0.1	1
SL	36	TVS line exit	LIQ/VAP	0.1	1

SUPPLY MODULE PRESSURE MEASUREMENTS

MEAS	NUM	DESCRIPTION	RANGE PSIA	INACCURACY +/- PSIA	SAMPLE FREQ (HZ)	MEAS TYPE
SP	1	Pressurization line press to supply tank	50	0.23	0.1 EDU	Control
SP	2	Supply vent line Pressure	50	0.23	0.1	Control EDU
SP	3	Relief vent line pressure	50	0.23	0.1	Control EDU
SP	4	Active TVS line pressure to mixer	50	0.37	0.1	Data
SP	5	Active TVS inlet pressure to supply tank	50	0.37	0.1	Data
SP	6	TVS inlet pressure to visco jets	50	0.37	0.1	Data
SP	7	Passive TVS exit line pressure to supply tank	50	0.37	0.1	Data
SP	8	Supply tank LH2 transfer line pressure	50	0.37	0.1	Data
SP	9	Supply tank fill drain line pressure	50	0.37	0.1	Data
SP	10	LH2 transfer line pressure downsteam of venturi bank	50	0.37	0.1	Data
SP	11	LH2 transfer line pressure to vaporizer	50	0.37	0.1	Data
SP	12	Vaporizer B coil inlet pressure	50	0.37	0.1	Data
SP	13	Vaporizer B coil exit pressure	200- 2000	37	0.1	Control
SP	14	Vaporizer A coil exit pressure	200- 2000	37	0.1	Control
SP	15	Vaporizer B bottle pressure	200- 2000	37	0.1	Control
SP	16	Vaporizer A bottle pressure	200- 2000	37	0.1	Control
SP	17	Vaporizer vent line pressure	50	0.37	0.1	Data
SP	18	Helium bottle pressure	3000	37	0.1	Data safety
SP	19	Helium bottle pressure	3000	37	0.1	Data safety

MEAS	NUM	DESCRIPTION	RANGE PSIA	INACCURACY +/- PSIA	SAMPLE FREQ (HZ)	MEAS TYPE
SP	20	Pressure at Venturi flow meter bank	50	0.37	0.1	Data
SP.	21	Supply tank active TVS flow pressure	25	0.18	0.1	Data
SP	22	Supply tank passive TVS flow pressure	25	0.18	0.1	Data
SP	23	Pressurant gas flow meter pressure	50	0.37	0.1	Data
SP	24	Flight Balanced vent pressure	50	0.37	0.1	Data
LARGE	RECEIVE	R (LR) TEMPERATURE M	EASUREMEN	rs		
MEAS	NUM	DESCRIPTION	RANGE Deg (R)	INACCURACY ± Deg (R)	SAMPLE FREQ. (Hz)	PRI. EXP.
LRTa	1	1 Pressurization line diffuser Temperature	36-540	< 1.5	0.1	8
LRTa	2	2 Internal vent line Temperature	29-46	< 0.2	0.1	3
LRTa	3	Top LAD Temperature	29-46	< 0.2	0.1	4
LRTa	4	Bottom LAD temp.	29-46	< 0.2	0.1	4
LRTa	5	Inside LAD channel #1, 3/4 length	29-46	< 0.2	0.1	4
LRTa	6	Inside LAD channel #2, 2/4 length	29-46	< 0.2	0.1	4
LRTa	7	Inside LAD channel #1, 1/2 length	29-46	< 0.2	0.1	4
LRTa	8	Inside LAD channel #2, 1/2 length	29-46	< 0.2	0.1	4
LRTa	9	Inside LAD channel #1, 1/4 length	29-46	< 0.2	0.1	4
LRTa	10	Inside LAD channel #2, 1/4 length	29-46	< 0.2	0.1	4
LRTa	11	Outside LAD channel #1, 3/4 length	29-46	< 0.2	0.1	4
LRTa	12	Outside LAD channel #2, 3/4 length	29-46	< 0.2	0.1	4
LRTa	13	Outside LAD channel #1, 1/2 length	29-46	< 0.2	0.1	4
LRTa	14	Outside LAD channel #2, 1/2 length	29-46	< 0.2	0.1	4

MEAS	NUM	DESCRIPTION	RANGE Deg (R)	INACCURACY ± Deg (R)	SAMPLE FREQ. (Hz)	PRI. EXP.
LRTa	15	Outside LAD channel #1, 1/4 length	29-46	< 0.2	0.1	4
LRTa	16	Outside LAD channel #2, 1/4 length	29-46	< 0.2	0.1	4
LRTa	17	Pressurization line diffuser temperature	29-46	< 0.2	0.1	8
LRTa	18	95% volume level Rake #1	29-46	< 0.2	0.1	3
LRTa	19	90% volume level Rake #2	29-46	< 0.2	0.1	3
LRTa	20	85% volume level Rake #1	29-46	< 0.2	0.1	3
LRTa	21	80% volume level Rake #2	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	22	75% volume level Rake #1	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	23	70% volume level Rake #2	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	24	65% volume level Rake #1	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	25	60% volume level Rake #2	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	26	55% volume level Rake #1	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	27	50% volume level Rake #3	29 - 46	< 1.0	0.1	3
LRTa	28	45% volume level Rake #5	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	29	40% volume level Rake #4	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	30	35% volume level Rake #5	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	31	30% volume level Rake #4	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	32	25% volume level Rake #5	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	33	20% volume level Rake #4	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	34	15% volume level Rake #5	36-110 110-540	< 1.0 < 1.5	0.1	3

MEAS	NUM	DESCRIPTION	RANGE Deg (R)	INACCURACY ± Deg (R)	SAMPLE FREQ. (Hz)	PRI. EXP.
LRTa	35	10% volume level Rake #4	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	36	5% volume level Rake #5	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	101	Top axial spray line temp. at tank wall	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	102	LH2 Transfer line temp. at tank wall	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	103	Lower axial spray line temp. at tank wall	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	104	Tangential spray line temp. at tank wall	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	105	vent line temperature at outer tank wall	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	106	Pressurization line temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	107	TVS line temp. at outer tank wall temp.	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	108	TVS line temperature at 50% length	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	109	TVS line temperature at 100% length	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	110	top dome warm side tank temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	111	top dome cold side tank temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	112	Aft dome warm side tank temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	113	Aft dome cold side tank temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	114	Upper barrel warm si tank temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	115	Upper barrel cold si tank temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	116	Central barrel, tank temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	117	Lower barrel warm si tank temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	118	Lower barrel cold si tank temperature	36-110 110-540	< 1.0 < 1.5	0.1	3

MEAS	NUM	DESCRIPTION	RANGE Deg (R)	INACCURACY ± Deg (R)	SAMPLE FREQ. (Hz)	PRI. EXP.
LRTa	119	Central barrel, tank temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	120	Warm side strut temperature at tank	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	121	Cold side strut temperature at tank	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	122	Warm side harness temperature at tank	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	123	Cold side harness temperature at tank	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	201	Top axial spray line temp. at panel L	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	202	LH2 Transfer line temp. at panel L	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	203	Lower axial spray line temp. at valve panel K	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	204	Tangential spray line temp. at valve panel K	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	205	Vent line temperature at valve panel L	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	206	Pressurization line temp. at panel L	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTa	207	TVS line temperature at valve panel L	36-110 110-540	< 1.0 < 1.5	0.1	3
LRTb	208	Forward dome honeycomb warm side	115-135	< 1.5	0.1	3
LRTb	209	Forward dome honeycomb cold side	115-135	< 1.5	0.1	3
LRTb	210	Aft dome, honeycomb warm side	115-135	< 1.5	0.1	3
LRTb	211	Aft dome, honeycomb cold side	115-135	< 1.5	0.1	3
LRTb	212	Honeycomb upper barr warm side	115-135	< 1.5	0.1	3
LRTb	213	Honeycomb lower barr cold side	115-135	< 1.5	0.1	3
LRTb	214	Honeycomb central temperature	115-135	< 1.5	0.1	3
LRTb	215	Honeycomb central temperature	115-135	< 1.5	0.1	3

MEAS	NUM	DESCRIPTION	RANGE Deg (R)	INACCURACY ± Deg (R)	SAMPLE FREQ. (Hz)	PRI. EXP.
LRTb	216	Warm side strut temperature	115-135	< 1.5	0.1	3
LRTb	217	Cold side strut temperature	115-135	< 1.5	0.1	3
LRTb	218	harness at honeycomb warm side temperature	115-135	< 1.5	0.1	3
LRTb	219	harness at honeycomb cold side	115-135	< 1.5	0.1	3
LRTb	301	Forward dome outer M temp. on warm side	360-520	< 1.5	0.1	3
LRTb	302	Aft dome cold side MLI temperature	230-520	< 1.5	0.1	3
LRTb	303	Aft barrel outer MLI temp. on warm side	360-520	< 1.5	0.1	3
LRTb	304	Forward barrel outer MLI temp. cold side	230-520	< 1.5	0.1	3
LRTb	305	MLI temp. central side +y direction	230-520	< 1.5	0.1	3
LRTb	306	MLI temp. central side tempy direct.	230-520	< 1.5	0.1	3
LRTa	307	Receiver tank flow transfer orifice temp	29 - 46	< 0.2	0.1	3
LRTb	308	Lg. Rec TVS flow meter inlet temp.	235-525	< 1.5	0.1	3
LARGE	RECEIVE	R LIQUID VAPOR POINT	MEASUREME	NTS		
MEAS	NUM	DESCRIPTION	RESPONSE		SAMPLE FREQ. (Hz)	PRI. EXP.
LRL	1	Rake #1 98% Volume	LIQ/VAP		0.1	2
LRL	2	Rake #2 96% Volume	LIQ/VAP		0.1	2
LRL	3	Rake #1 94% Volume	LIQ/VAP		0.1	2
LRĹ	4	Rake #2 92% Volume	LIQ/VAP		0.1	2
LRL	5	Rake #1 90% Volume	LIQ/VAP		0.1	2
LRL	6	Rake #2 88% Volume	LIQ/VAP		0.1	2
LRL	7	Rake #1 86% Volume	LIQ/VAP		0.1	2
LRL	8	Rake #2 84% Volume	LIQ/VAP		0.1	2
LRL	9	Rake #1 82% Volume	LIQ/VAP		0.1	2
LRL	10	Rake #2 80% Volume	LIQ/VAP		0.1	2

MEAS	NUM	DESCRIPTION	RESPONSE	SAMPLE FREQ. (Hz)	PRI. EXP.
LRL	11	Rake #1 78% Volume	LIQ/VAP	0.1	2
LRL	12	Rake #2 76% Volume	LIQ/VAP	0.1	2
LRL	13	Rake #1 74% Volume	LIQ/VAP	0.1	2
LRL	14	Rake #2 72% Volume	LIQ/VAP	0.1	2
LRL	15	Rake #1 70% Volume	LIQ/VAP	0.1	2
LRL	16	Rake #2 66% Volume	LIQ/VAP	0.1	2
LRL	17	Rake #1 62% Volume	LIQ/VAP	0.1	2
LRL	18	Rake #2 58% Volume	LIQ/VAP	0.1	2
LRL	19	Rake #1 54% Volume	LIQ/VAP	0.1	2
LRL	20	Rake #3 50% Volume	LIQ/VAP	0.1	2
LRL	21	Rake #5 46% Volume	LIQ/VAP	0.1	2
LRL	22	Rake #4 42% Volume	LIQ/VAP	0.1	2
LRL	23	Rake #5 38% Volume	LIQ/VAP	0.1	2
LRL	24	Rake #4 34% Volume	LIQ/VAP	0.1	2
LRL	25	Rake #5 30% Volume	LIQ/VAP	0.1	2
LRL	26 -	Rake #4 26% Volume	LIQ/VAP	0.1	2
LRL	27	Rake #5 22% Volume	LIQ/VAP	0.1	2
LRL	28	Rake #4 18% Volume	LIQ/VAP	0.1	2
LRL	29	Rake #5 14% Volume	LIQ/VAP	0.1	2
LRL	30	Rake #4 10% Volume	LIQ/VAP	0.1	2
LRL	31	Rake #5 6% Volume	LIQ/VAP	0.1	2
LRL	32	Vent line exit	LIQ/VAP	0.1	2
LRL	33	Upper LAD box	LIQ/VAP	0.1	4
LRL	34	LAD channel #1 3/4 length	LIQ/VAP	0.1	4
LRL	35	LAD channel #2 3/4 length	LIQ/VAP	0.1	4
LRL	36	LAD channel #1 1/2 length	LIQ/VAP	0.1	4
LRL	37	LAD channel #2 1/2 length	LIQ/VAP	0.1	4
LRL	38	LAD channel #1 1/4 length	LIQ/VAP	0.1	4

MEAS	NUM	DESCRIPTION	RESPONSE		SAMPLE FREQ. (Hz)	PRI. EXP.
LRL	39	LAD channel #2 1/4 length	LIQ/VAP		0.1	4
LRL	40	Lower LAD box	LIQ/VAP		0.1	4
LARGE	RECEIVE	R PRESSURE MEASUREM	ENTS			
MEAS	NUM	DESCRIPTION	RANGE PSIA	ACCURACY +/- PSIA	SAMPLE FREQ. (Hz)	MEAS TYPE
LR	1	Lg Rec relief vent meter pressure	50	0.23	0.1	Control EDU
LR	2	Lg Rec pressuriza- tion line pressure	50	0.23	0.1	Control EDU
LR	3	Lg. Rec vent line pressure	50	0.23	0.1	Control EDU
LR	4	Top axial spray line pressure	50	0.37	0.1	Data
LR	5	Lg Rec LAD transfer line pressure	50	0.37	0.1	Data
LR	6	Aft axial spray line pressure	50	0.37	0.1	Data
LR	7	Tangential spray line pressure	50	0.37	0.1	Data
LR	8	Receiver tanks venturi pressure	50	0.37	0.1	Data
LR	9	Lg. Rec TVS line pressure	25	0.18	0.1	Data
SMALL	RECEIVE	R (SR) MODULE TEMPER	ATURE MEASUR	REMENTS		
MEAS	NUM	DESCRIPTION	RANGE Deg (R)	INACCURACY ± Deg (R)	SAMPLE RATE (Hz)	PRI. EXP.
SRTa	1	Pressurization line diffuser temp.	36-110 110-540	< 1.0	0.1	3
SRTa	2	Vent line exit temperature	29 - 46	< 0.2	0.1	3
SRTa	3	97.5% volume level radial spray bar	29 - 46	< 0.2	0.1	3
SRTa	4	80.0% volume level radial spray bar	29 - 46	< 0.2	0.1	3
SRTa	5	50.0% volume level radial spray bar	29 - 46	< 0.2	0.1	3
SRTa	6	20.0% volume level radial spray bar	29 - 46	< 0.2	0.1	3 ,

MEAS	NUM	DESCRIPTION	RANGE Deg (R)	INACCURACY ± Deg (R)	SAMPLE RATE (Hz)	PRI. EXP.
SRTa	7	2.5% volume level radial spray bar	29 - 46	< 0.2	0.1	3
SRTa	8	Axial spray line temperature	29 - 46	< 0.2	0.1	3
SRTa	9	Tangential spray line left nozzle temp.	29 - 46	< 0.2	0.1	3
SRTa	10	Tangential spray line right nozzle temp.	29 - 46	< 0.2	0.1	3
SRTa	11	Tangential spray line center left temp.	29 - 46	< 0.2	0.1	3
SRTa	12	Tangential spray line center right temp.	29 - 46	< 0.2	0.1	3
SRTa	13	Rake V-1, 5% level	29 - 46	< 0.2	0.1	3
SRTa	14	Rake V-1, 75% level	29 - 46	< 0.2	0.1	3
SRTa	15	Rake V-1, 75% level	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	16	Rake V-1, 95% level	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	17	Rake V-2, 5% level	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	18	Rake V-2, 25% level	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	19	Rake V-2, 75% level	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	20	Rake V-2, 95% level	29 - 46	< 0.2	0.1	3
SRTa	21	Rake H-2 center right temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	22	Rake H-2 center left temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	23	Rake H-1 center right temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	24	Rake H-1 center left temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
SMALL	RECEIVE	ER TANK OUTER WALL	TEMPERATURE N	MEASUREMENTS		
MEAS	NUM	DESCRIPTION	RANGE Deg (R)	ACCURACY ± Deg (R)	SAMPLE RATE (Hz)	PRI. EXP.
SRTa	101	Top axial spray line temp. at tank wall	36-110 110-540	< 1.0 < 1.5	0.1	3

MEAS	NUM	DESCRIPTION	RANGE Deg (R)	ACCURACY ± Deg (R)	SAMPLE RATE (Hz)	PRI. EXP.
SRTa	102	Radial spray line temp. at tank wall	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	103	Tangential spray line temp. at tank wall	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	104	Vent line temp. at outer tank wall	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	105	Pressurization line temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	106	TVS line temperature downstream of JT device	29 - 46	< 0.2	0.1	3
SRTa	107	TVS line temperature 50% length	29 - 46	< 0.2	0.1	3
SRTa	108	TVS line temperature 100% length	29 - 46	< 0.2	0.1	3
SRTa	109	Top dome, quadrant 1 temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	110	Top dome, quadrant 2 temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	111	Top dome, quadrant 3 temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	112 .	Aft dome, quadrant 4 temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	113	Aft dome, quadrant 5 temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	114	Aft dome, quadrant 6 temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	115	Tangential spray are tank temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	116	Tangential spray are tank temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	117	Barrel Section 90 de from spary	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	118	Barrel Section 90 de from spary	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	119	Warm side harness temperature at tank	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	120	Cold side harness temperature at tank	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	121	Warm side strut temperature at tank	36-110 110-540	< 1.0 < 1.5	0.1	3

MEAS	NUM	DESCRIPTION	RANGE Deg (R)	ACCURACY ± Deg (R)	SAMPLE RATE (Hz)	PRI. EXP.
SRTa	122	Cold side strut temperature at tank	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	201	LH2 Transfer line temp. at panel M	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	202	Lower axial spray line temp. at valve panel M	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	203	Radial spray line temperature at panel M	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	204	Tangential spray line temp. at panel N valve	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	205	vent line temperature at valve panel N	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTa	206	Pressurization line temperature at panel N	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTb	207	Top dome, honeycomb quadrant 1	117-135	< 1.5	0.1	3
SRTb	208	Top dome, honeycomb quadrant 2	117-135	< 1.5	0.1	3
SRTb	209	Top dome, honeycomb quadrant 3	117-135	< 1.5	0.1	3
SRTb	210	Aft dome, honeycomb quadrant 4	117-135	< 1.5	0.1	3
SRTb	211	Aft dome, honeycomb quadrant 5	117-135	< 1.5	0.1	3
SRTb	212	Aft dome, honeycomb quadrant 6	117-135	< 1.5	0.1	3
SRTb	213	Honeycomb barrel temperature +y	117-135	< 1.5	0.1	3
SRTb	214	Honeycomb barrel temperature -y	117-135	< 1.5	0.1	3
SRTb	215	Honeycomb barrel temperature +z	117-135	< 1.5	0.1	3
SRTb	216	Honeycomb barrel temperature +z	117-135	< 1.5	0.1	3
SRTb	217	Warm side harness temperature at IMLI	117-135	< 1.5	0.1	3
SRTb	218	Cold side harness temperature at IMLI	117-135	< 1.5	0.1	3
SRTb	219	Warm side strut temperature at IMLI	117-135	< 1.5	0.1	3

MEAS	NUM	DESCRIPTION	RANGE Deg (R)	ACCURACY ± Deg (R)	SAMPLE RATE (Hz)	PRI. EXP.
SRTb	220	Cold side strut temperature at IMLI	117-135	< 1.5	0.1	3
SRTa	301	Dump vent line temperature	36-110 110-540	< 1.0 < 1.5	0.1	3
SRTb	302	Top dome, outer MLI quadrant 1	235-525	< 1.5	0.1	3
SRTb	303	Top dome, outer MLI quadrant 2	235-525	< 1.5	0.1	3
SRTb	304	Aft doem, outer MLI quadrant 4	235-525	< 1.5	0.1	3
SRTb	305	Aft doem, outer MLI quadrant 5	235-525	< 1.5	0.1	3
SRTb	306	Outer MLI barrel temperature	235-525	< 1.5	0.1	3
SRТъ	307	Outer MLI barrel temperature	235-525	< 1.5	0.1	3
SRTb	309	Outer MLI barrel temperature	235-525	< 1.5	0.1	3
SRTb	310	Sm. Rec TVs flow meter inlet temperature	235-525	< 1.5	0.1	3
SMALL	RECEIVE	R TANK MODULE LIQUII	VAPOR POINT	MEASUREMENTS		
MEAS	RECEIVE NUM	R TANK MODULE LIQUIL DESCRIPTION	PO VAPOR POINT I RESPONSE Deg (R)	MEASUREMENTS	SAMPLE FREQ. (Hz)	PRI. EXP.
			RESPONSE	MEASUREMENTS	SAMPLE	
MEAS	NUM	DESCRIPTION Radial Spray Rake	RESPONSE Deg (R)	MEASUREMENTS	SAMPLE FREQ. (Hz)	EXP.
MEAS SRL	NUM I	DESCRIPTION Radial Spray Rake 95.5% level Rake H-1, -8.2	RESPONSE Deg (R) LIQ/VAP	MEASUREMENTS	SAMPLE FREQ. (Hz) 0.1	EXP.
MEAS SRL SRL	NUM I 2	DESCRIPTION Radial Spray Rake 95.5% level Rake H-1, -8.2 inch from center Rake H-1, -4.1	RESPONSE Deg (R) LIQ/VAP	MEASUREMENTS	SAMPLE FREQ. (Hz) 0.1	EXP. 2
MEAS SRL SRL SRL	NUM I 2 3	DESCRIPTION Radial Spray Rake 95.5% level Rake H-1, -8.2 inch from center Rake H-1, -4.1 inch from center	RESPONSE Deg (R) LIQ/VAP LIQ/VAP LIQ/VAP	MEASUREMENTS	SAMPLE FREQ. (Hz) 0.1 0.1	2 2 2
MEAS SRL SRL SRL	NUM 1 2 3	DESCRIPTION Radial Spray Rake 95.5% level Rake H-1, -8.2 inch from center Rake H-1, -4.1 inch from center Rake H-1, center Rake H-1, center	RESPONSE Deg (R) LIQ/VAP LIQ/VAP LIQ/VAP	MEASUREMENTS	SAMPLE FREQ. (Hz) 0.1 0.1 0.1	2 2 2 2
MEAS SRL SRL SRL SRL SRL	NUM 1 2 3 4 5	DESCRIPTION Radial Spray Rake 95.5% level Rake H-1, -8.2 inch from center Rake H-1, -4.1 inch from center Rake H-1, center Rake H-1, 8.2 inch from center Rake H-1, 8.2 inch from center	RESPONSE Deg (R) LIQ/VAP LIQ/VAP LIQ/VAP LIQ/VAP	MEASUREMENTS	SAMPLE FREQ. (Hz) 0.1 0.1 0.1 0.1	2 2 2 2 2
MEAS SRL SRL SRL SRL SRL	NUM 1 2 3 4 5	DESCRIPTION Radial Spray Rake 95.5% level Rake H-1, -8.2 inch from center Rake H-1, -4.1 inch from center Rake H-1, center Rake H-1, 8.2 inch from center Rake H-1, 4.1 inch from center Rake H-1, 4.1 inch from center	RESPONSE Deg (R) LIQ/VAP LIQ/VAP LIQ/VAP LIQ/VAP LIQ/VAP LIQ/VAP	MEASUREMENTS	SAMPLE FREQ. (Hz) 0.1 0.1 0.1 0.1 0.1	2 2 2 2 2 2
MEAS SRL SRL SRL SRL SRL SRL	NUM 1 2 3 4 5 6	Radial Spray Rake 95.5% level Rake H-1, -8.2 inch from center Rake H-1, -4.1 inch from center Rake H-1, center Rake H-1, 8.2 inch from center Rake H-1, 8.1 inch from center Rake H-1, 4.1 inch from center Radial Spray Rake 92.5% level	RESPONSE Deg (R) LIQ/VAP LIQ/VAP LIQ/VAP LIQ/VAP LIQ/VAP LIQ/VAP	MEASUREMENTS	SAMPLE FREQ. (Hz) 0.1 0.1 0.1 0.1 0.1 0.1 0.1	2 2 2 2 2 2

MEAS	NUM	DESCRIPTION	RESPONSE Deg (R)	SAMPLE FREQ. (Hz)	PRI. EXP.
SRL	11	Rake V-1, 84% level	LIQ/VAP	0.1	2
SRL	12	Radial bar, 82%	LIQ/VAP	0.1	2
SRL	13	Rake V-2, 80% level	LIQ/VAP	0.1	2
SRL	14	Rake V-1, 78% level	LIQ/VAP	0.1	2
SRL	15	Radial bar, 76%	LIQ/VAP	0.1	2
SRL	16	Rake V-2, 74% level	LIQ/VAP	0.1	2
SRL	17	Rake V-1, 72% level	LIQ/VAP	0.1	2
SRL	18	Radial bar, 70%	LIQ/VAP	0.1	2
SRL	19	Rake V-1, 65% level	LIQ/VAP	0.1	2
SRL	20	Radial bar, 60%	LIQ/VAP	0.1	2
SRL	21	Rake V-2, 55% level	LIQ/VAP	0.1	2
SRL	22	Tangential Rake -13.3 in from center	LIQ/VAP	0.1	2
SRL	23	Tangential Rake -9.7 in from center	LIQ/VAP	0.1	2
SRL	24	Tangential Rake center	LIQ/VAP	0.1	2
SRL	25	Tangential Rake 13.3 in from center	LIQ/VAP	0.1	2
SRL	26	Tangential Rake 9.7 in from center	LIQ/VAP	0.1	2
SRL	27	Rake V-1, 45% level	LIQ/VAP	0.1	2
SRL	28	Radial bar, 40%	LIQ/VAP	0.1	2
SRL	29	Rake, V-2, 35% level	LIQ/VAP	0.1	2
SRL	30	Rake V-1, 30% level	LIQ/VAP	0.1	2
SRL	31	Radial bar, 25%	LIQ/VAP	0.1	2
SRL	32	Rake V-2, 20% level	LIQ/VAP	0.1	2
SRL	33	Rake V-1, 15% level	LIQ/VAP	0.1	2
SRL	34	Rake V-2, 10% level	LIQ/VAP	0.1	2
SRL	35	Rake H-2, -8.2 inch from center	LIQ/VAP	0.1	2
SRL	36	Rake H-2, -4.1 inch from center	LIQ/VAP	0.1	2
SRL	37	Rake H-2, center	LIQ/VAP	0.1	2

MEAS	NUM	DESCRIPTION	RESPONSE Deg (R)		SAMPLE FREQ. (Hz)	PRI. EXP.
SRL	38	Rake H-2, 8.2 inch from center	LIQ/VAP		0.1	2
SRL	39	Rake H-2, 4.1 inch from center	LIQ/VAP		0.1	2
SRL	40	Vent eixt line	LIQ/VAP		0.1	2
SMALL	RECEIVE	R MODULE PRESSURE M	EASUREMNTS			
MEAS	NUM	DESCRIPTION	RANGE PSIA	ACCURACY +/- PSIA	SAMPLE RATE (HZ)	MEAS TYPE
SRP	1	Sm Rec pressurization line pressure	50	0.23	0.1	Control EDU
SRP	2	Sm. Rec vent line pressure	50	0.23	0.1	Control EDU
SRP	3	Sm Rec dump vent line pressure	50	0.23	0.1	Control EDU
SRP	4	LH2 transfer line pressure	50	0.37	0.1	Data
SRP	5	axial spray line pressure	50	0.37	0.1	Data
SRP	6	radial spray line pressure	50	0.37	0.1	Data
SRP	7	Tangential spray line pressure	50	0.37	0.1	Data
SRP	8	Pressure downstream of JT device	50	0.37	0.1	Data
SRP	9	Sm Rec TVS flow line pressure	25	0.18	0.1	Data
COLD-S	AT EXPE	RIMENT SUBSYSTEM AC	CELERATION M	EASUREMENTS		
MEAS	NUM	DESCRIPTION	RANGE (MICRO-G)	ACCURACY RATE (Hz)	SAMPLE	PRI EXP
SA	1	X-axis acceleration	1-100	TBD	10	3
SA	2	Y-axis acceleration	1-100	TBD	10	3
SA	3	Z-axis acceleration	1-100	TBD	10	3
COLD-S	AT EXPE	RIMENT SUBSYSTEM PO	WER (W) MEASU	JREMENTS		=
MEAS	NUM	DESCRIPTION	RANGE WATTS	ACCURACY	SAMPLE RATE (Hz)	PRI. EXP.
w	1	Supply tank heater power	8-21	TBD	1	1

MEAS	NUM	DESCRIPTION	RANGE WATTS	ACCURACY	SAMPLE RATE (Hz)	PRI. EXP.
w	2	Vaporizer A	100	TBD	1	Eng
w	3	Vaporizer B	100	TBD	1	Eng
w	4	Large receiver heater power	50	ТВО	1	3
w	5	Small reciever heater power	50	ТВО	1	3
\mathbf{w}	6	Flight vent	7	TBD	1	3
w	7	TVS vent	7	TBD	1	1
\mathbf{w}	8	Vaporizer panel A	6	TBD	Ī	Eng
w	9	Vaporizer panel B	6	TBD	1	Eng
w	10	Connector panel A	2	TBD	1	Eng
w	11	Connector panel B	2	TBD	1	Eng
COLD-	SAT EXPE	RIMENTAL SUBSYSTEM	FLOW MEASUE	REMENTS		
MEAS	NUM	DESCRIPTION	RANGE (LBM/HR)	INACCURACY FREQ.	SAMPLE	PRI EXP
F	1	LH2 transfer flow from supply tank	50-350	0.5% read	TBD	3
F	2	LH2 low flow rate	15-50	0.5% span	TBD	3
F	3	LH2 medium flow rate	30-100	0.5% span	TBD	3
F	4	LH2 high flow rate	60-200	0.5% span	TBD	3
F	5	Supply tank passive TVS flow rate	0.16	1.0% span	TBD	1
F	6	Supply tank active TVS flow rate	5.6	1.0% span	TBD	1
F	7	Lg. Rec. tank passive TVS flow rate	0.21	1.0% span	TBD	3
F	8	Sm. Rec. tank passive TVS flow rate	0.15	1.0% span	TBD	3
F	9	Pressurant flow rate	1-10	1.0% span	TBD	3
F	10	LH2 transfer flow between receivers	60-200	0.5% span	TBD	3
F	11	Flight balanced Vent flow rate	TBD	TBD	TBD	8

TABLE 1.— EXPERIMENT ELECTRICAL AND ELECTRONICS BOX DEFINITION

			47.45	
Type of measurement unit	Number of units required	Electronics bay location	Unit weight, lb	Unit power, W
Experimental data	3	2	7	15
Accelerometer data	1	2	6	10
Mixer motor power	1	2	7.5	15
Capacitive probe signal conditioning power	I	2	12	15

TABLE 2.—8-BIT DATA ACQUISITION
SYSTEM VOLTAGE MEASUREMENT
UNCERTAINTY ESTIMATES

Voltage range, mV	Uncertainty, mV	rss uncertainty, percent
0-10	±0.18	1.80
0-30	±.20	.67
0-125	±.45	.36
0-1250	+3 34	27

TABLE 3.—COLD-SAT EXPERIMENT SUBSYSTEM MEASUREMENT REQUIREMENT SUMMARY

 _			-			
Measurement requirement	Measurement range	Transistor type	Measurement error	Resolution, bit	Sample frequency, Hz	Number required
High-accuracy fluid temperature	30.0–50.0 °R	Type A PRT	≤±0.2 R	8	0.1	90
Structure temperature	30.0-50.0 °R	Type A PRT	≤±0.2 R	8	0.1	49
Structure temperatures	36.0–540 °R	Type AB PRT	≤±2.0 R	8	0.1	162
Hydrogen liquid/vapor level detection	Tank volume	Thermistor	±1.0 percent	1	0.1	106
Two-phase flow detection	Liquid to vapor	Thermistor	NA	1	0.1	10
High resolution tank pressures	0-50 psia	Cryogenic strain gauge	±0.23 psia	12	0.1	9
Plumbing system pessures	0-5000 psia	Cryogenic strain gauge	±1.0%	8	0.1	33
LH2 Transfer Flow Rates	50, 100, 200 lbm/hr	Venturi/ΔP	<±2.0%	12	0.1	4
Mixer flow rates	TBD	rpm	TBD	12	0.1	2
TVS flow rates: Supply tank Large receiver Small receiver	lbm/hr 0.16–5.6 0.21 0.15	Turbine meters	<±5.0%	12 12 12	0.1 .1 .1	2 1 1
Vent flow rates	5-50 lbm/hr	Sonic flow nozzle	<±5.0%	8	10	3
Acceleration	≤ 100 μg	Accelerometer	TBD	12	10	3
Valve status Open/close Switch		Switch	N/A	1	0.1	61

TABLE 4.—TEMPERATURE SENSOR CANDIDATE REQUIREMENTS AND ERRORS

Sensor type	Temperature range, °R	Excitation level, mA	Excitation error, percent	Sensor output, mV	Total 8-bit measurement error, mV
GRT	20–50	(a)	±0.1	≤10	0.18
PRT R _o = 1000	29-47	10	±0.1	33–125	0.45
	29–540	1.0	±0.1	3.3–125 125–1250	0.45 3.3
Silicon diode	20-540	0.01	±0.1	1300-520	3.3

^{*}Variable.

TABLE 5.—ERROR ANALYSIS FOR 1000-OHM PLATINUM
RESISTANCE THERMOMETER IN 8-BIT DATA
ACQUITISION SYSTEM

[DAS error, ±0.45 mV; excitation current, 10 mA.]

Temper- ature, °R	Resis- tance, Ω	Sensivity of resistance to temperature, ΔΩ/Δ°R	Sensitiv- ity of temper- ature to voltage, °R/mV	Temper- ature error °R	Error, percent
29.07 3.87 32.67 34.47 36.27 38.07 39.87 41.67 50.67	3.39 3.84 4.37 4.99 5.70 6.51 7.42 8.45 15.33	0.23 .27 .32 .37 .42 .48 .54 .73	0.43 .37 .31 .27 .24 .21 .19 .14	±0.20 ±.17 ±.14 ±.12 ±.11 ±.10 ±.08 ±.06 ±.05	±0.68 ±.54 ±.44 ±.36 ±.30 ±.25 ±.21 ±.15 ±.10

TABLE 6.—8-BIT DATA ACQUISITION SYSTEM INACCURACY INFLUENCE ON GRT TEMPERATURE MEASUREMENT ERROR

[Voltage measurement error, ±0.18 mV.]

Temper- ature, °R	Resis- tance, Ω	Sensitiv- ity of resistance to temper- ature, Ω/°R	Current, mA	Sensitiv- ity of temper- ature to voltage, °R/mV	Temper- ature error, percent	Error, percent
18	149.16	-21.89	0.067	0.681	±0.1	±0.67
36	23.57	-1.58	.424	1.490	±.27	±.75
54	9.46	36	1.058	2.647	±.48	±.89
72	5.44	13	1.839	4.130	±.74	±1.0
180	1.695	007	5.900	25.45	±4.6	±2.6

TABLE 7.—UNCERTAINTY IN SATURATION TEMPERATURE DUE TO PRESSURE MEAS-UREMENT UNCERTAINTY (UP)

[Uncertainty in saturation pressure, UP, ±0.23 psi.]

Temper- ature, °R	Pressure, psia	Sensitivity of temperature to pressure, $\delta T/\delta P$, °R/psi	Uncertainty in saturation temperature, UT, °R
30.806 36.608 41.299 45.406	5 15 30 50	0.7836 .4231 .2431 .1385	±0.18 ±0.10 ±0.06 ±0.03

TABLE 8.—FLIGHT QUALIFIED CRYOGENIC PRESSURE SENSOR CANDIDATE SPECIFICATIONS

I	_
١	Transducer type Cryogenic strain gauge
	Temperature range, °R
	Full scale inaccuracy, percent ±0.25
1	Thermal zero shift, percent of FS/°R ±0.015
	Thermal span shift, percent/°R ±0.005
	Natural frequency (at 15 psi), kHz
	Thermal span shift, percent/°R

Table 9.—PRESSURE MEASUREMENT INACCURACY

AND RESOLUTION VALUES

Full scale pressure, psia	8-bit system error, psia	8-bit resolution, psi	12-bit system error, psia	12-bit resolution, psi
15	±0.11	0.06	±0.07	0.004
25	±.18	.10	±.11	.006
30	±.21	.12	±.13	.007
50	±.36	.20	±.23	.012

TABLE 10.—LIQUID HYDROGEN FLOW MEASUREMENT ERROR ANALYSIS

[Calibration, ±0.5 percent; density, ±0.2 percent.]

(a) Orifice differential pressure

	Flow range, percent	ΔP error, percent	EDU error, percent	Mass flow rate rss error, percent
	100	±0.25	±0.20	±0.63
1	90	±.31	±.25	±.67
	80	±.39	±.31	±.74
1	70	±.51	±.41	±.85
	60	±.69	±.56	±1.04
	50	±1.00	±.80	±1.39
	40	±1.56	±1.25	±2.07

(b) Turbine flowmeter

Flow range, percent	EDU error, percent	Mass flow rate rss error, percent
100 90 80 70 60 50 40	±0.40 ±.44 ±.50 ±.57 ±.67 .80 ±1.00	±0.67 ±.72 ±.74 ±.79 ±.85 ±.96 ±1.16

TABLE 11.-THERMODYNAMIC VENT SYSTEM (TVS) FLOW RATE RANGE AND TURBINE METER SIZING

Location	TVS type	Mass flow	Meter inlet temper- ature and pressure		Turbine meter	Flow range,
		rate, lb/hr	180 °R, 20 psia	360 °R, 5 psia	diameter required, in.	ft³/hr
			Equivalent volume flow rate, ft ³ /min			
Small receiver Large receiver Supply tank Supply tank	Passive Passive Passive Active	0.15 .21 .16 5.6	0.12 .17 .13 4.47	0.96 1.34 1.02 35.84	0.5 .5 .5 1.0	0.1-1.0 0.25-2.5 0.1-1.0 5-50

TABLE 12.-TVS FLOW MEASUREMENT

ERROR ANALYSIS

Meter range, percent	Turbine inaccuracy, percent	EDU error, percent	Density error, percent	Mass flow rate rss error, percent
100	±1.05	±0.4	±3.66	±3.83
50	±2.10	±0.8	±3.66	±4.29

TABLE 13.-TEMPERATURE INFLUENCE ON NOZZLE DIAMETER

Temperature °R	Density, lb/ft ³	Velocity, V _s , ft/sec	Area, in. ²	Diameter, in.	Volume flow rate, ft ³ /min
50	0.222	1334	6.75×10^{-3} 1.28×10^{-2} 1.80×10^{-2} 2.16×10^{-2} 2.43×10^{-2}	0.093	3.75
140	.0675	2320		.128	12.37
240	.0391	2842		.151	21.32
340	.0276	3363		.166	30.2
440	.0213	3860		.176	39.13

TABLE 14.—LARGE RECEIVER ESTIMATED VENT TIMES

Temper-	Mass, lb	, at —	Mass difference,	Vent time,
ature °R			lb	564
440 340 240 140	0.267 .345 .489 .838	0.222 .288 .408 .698	0.045 .057 .081 .140	8 9 11 13

TABLE 15.—VENT FLOW RATE MEASUREMENT ERROR ANALYSIS RESULTS

Temper-	Pressure error, percent, at—								
ature, °R	ature, 50 psia		35 psia 25 psia		5 psia				
140 440	±1.01 ±.90	±1.12 ±1.01	±1.24 ±1.20	±1.78 ±1.71	±4.70 ±4.67				

TABLE 16.—SUPPLY MODULE INSTRUMENTATION LISTING

Sensor location				S	ensor type				
	Temper- ature sensors	Liquid- vapor sensors	Pressure sensors	Differ- ential pressure venturi	Turbine meters	Cryogen valve discretes	Gas valve discretes	Speed sensors	Capaci- tance probe
Inside pressure vessel	54	34	0	0	0	0	0	0	1
Liquid acquisition device	10	2	0	0	0	0	0	0	0
Outer tank wall	10	0	0	0	0	0	0	0	0
Tank wall plumbing	5	0	0	0	0	0	0	0	0
Tank wall struts	2	0	0	0	0	0	0	0	0
Tank wall harnesses	2	0	0	0	0	0	0	0	0
Panel E inside MLI	1	0	1	0	0	2	0	0	0
Panel F inside MLI	2	0	2	0	0	3	0	2	0
Panel G inside MLI	1	0	2	0	0	2	0	0	0
Panel H inside MLI	3	0	2	0	0	3	0	0	0
Panel I inside MLI	2	0	1	4	0	3	0	0	0
Panel J inside MLI	2	0	3	0	1	5	0	0	0
Struts inside MLI	2	0	0	0	0	0	0	0	0
Harnesses inside MLI	2	0	0	0	0	0	0	0	0
Honeycomb wall	10	0	0	0	0	0	0	0	0
Outside MLI can	10	0	0	0	0	0	0	0	0
Tray	I	0	1	0	1	4	0	0	0
Hydrogen vaporizers	10	0	7	0	0	4	8	0	0
Helium tanks	5	0	3	0	1	0	4	0	0
Vent panel O	2	0	2	0	1	0	0	0	0
Accelerometer	3	0	0	0	0	0	0	0	0
Total:	139	36	24	4	4	26	12	2	1

TABLE 17 - LARGE RECEIVER MODULE INSTRUMENTATION LISTING

Location			Senso	r type		
	Temper- ature sensors	Liquid- vapor sensors	Pressure sensors	Differ- ential pressure venturi	Turbine meters	Cryogen valve discretes
Inside Pressure vessel	22	32	0	0	0	0
Liquid acquisition devise	14	8	0	0	0	0
Outer tank wall	10	0	0	0	0	0
Tank wall plumbing & TVS	9	0	0	0	0	0
Tank wall struts	2	0	0	0	0	0
Tank wall harnesses	2	0	0	0	0	0
Panel K inside MLI	3	0	3	0	0	3
Panel L inside MLI	4	0	5	0	1	4
Struts inside MLI	2	0	0	0	0	0
Harnesses inside MLI	2	0	0	0	0	0
Honeycomb wall	8	0	0	0	0	0
Outside MLI wall	6	0	0	0	0	0
Tray	2	0	1	2	0	2
Total:	86	40	9	2	1	9

TABLE 18.—SMALL RECEIVER MODULE INSTRUMENTATION LISTING

Location			Sens	or type		<u></u>
	Temper- ature sensors	Liquid- vapor sensors	Pres- sure sensors	Differ- ential pressure venturi	Turbine meters	Cryogen valve discretes
Inside pressure vessel	24	40	0	0	0	0
Outer tank wall	10	8	0	0	0	0
Tank wall plumbing & TVS	8	0	0	0	0	0
Tank wall struts	2	0	0	0	0	0
Tank wall harnesses	2	0	0	0	0	0
Panel M inside MLI	3	0	4	0	0	4
Panel N inside MLI	3	0	4	0	1	4
Struts inside MLI	2	0	0	0	0	0
Harnesses inside MLI	2	0	0	0	0	0
Honeycomb wall	10	0	0	0	0	0
Outside MILI wall	8	0	0	0	0	0
Tray	2	0	1 :	0	0	2
Total:	86	40	9	2	1	9

TABLE 19.—ESTIMATED WIRE HARNESS WEIGHT

Wire size	Conductor	Insulation material	Wire diameter, in.	Estimated weight, lb/1000 ft
24 AWG	Manganin	Polyimide	0.037	2.0
18 AWG	Phosphor- bronze	Polyimide	.062	6.6

TABLE 20.— ESTIMATED WIRE

BUNDLE SIZE

Wire size	Conductor material	Number of wires	Estimated bundle diameter, in.
24 AWG	Manganin	45 50 55	0.284 .299 .313
18 AWG	Phosphor- bronze	40 35 45	0.451 .422 .477

TABLE 21.—COLD-SAT EXPERIMENT SUBSYSTEM WIRING SUMMARY

Module	Numb instrumenta		Numl power	Total wires	
	24-AWG manganin	24-AWG copper	18-AWG phorphor- bronze	24-AWG copper	
Supply module Large receiver module Small receiver module Total:	547 391 370 1308	188 0 0 188	66 40 44 152	160 0 0 160	963 431 414 1808

TABLE 22.—SUPPLY MODULE INSTRUMENTATION WIRE ALLOCATION

Location				Ser	sor or hard	lware type				
	Temper- ature wires	Liquid- vapor wires	Pres- sure wires	Differ- ential pressure wires	Turbine wires	Cryogen valve wires	Gas valve wires	Speed sensor wires	Capaci- tance probe	Total wires
Inside pressure vessel	164	21	0	0	0	0	0	0	4	189
Outer tank wall	40	0	0	0	0	0	0	0	0	40
Tank wall plumbing	20	0	0	0	0	0	0	0	0	20
Tank wall struts	8	0	0	0	0	0	0	0	0	8
Tank wall harnesses	8	0	0	0	0	0	0	0	0	8
Panel E inside MLI	4	0	4	0	0	6	0	0	0	14
Panel F inside MLI	8	0	8	0	0	9	0	4	0	29
Panel G inside MLI	4	. 0	8	0	0	6	0	0	0	18
Panel H inside MLI	12	0	8	0	0	9	0	0	0	29
Panel I inside MLI	8	0	4	16	0	9	0	0	0	37
Panel J inside MLI	8	0	12	0	2	15	0	0	0	37
Struts inside MLI	8	0	0	0	0	0	0	0	0	8
Harnesses inside MLI	8	0	0	0	0	0	0	0	0	8
Honeycomb wall	40	0	0	0	0	0	0	0	0	40
Outside MLI can	40	0	0	0	0	0	0	0	0	40
Tray	4	0	4	0	2	12	0	0	0	22
Hydrogen vaporizers (Cu)	20	0	28	0	0	12	48	0	0	108
Helium tanks (Cu)	10	0	12	0	2	0	24	0	0	48
Vent panel O (Cu)	4	0	8	0	2	12	0	0	0	26
Accelerometer (Cu)	6	0	0	0	0	0	0	0	0	6
Total:	424	21	96	16	8	90	72	4	4	735

TABLE 23.—LARGE RECEIVER MODULE INSTRUMENTATION WIRE ALLOCATIONS

Location				Sensor type			
	Temper- ature wires	Liquid- vapor wires	Pressure wires	Differ- ential pressure wires	Turbine meters	Cryogen valve wires	Total wires
Inside pressure vessel	96	22	0	0	0	0	118
Outer tank wall	40	0	0	0	0	0	40
Tank wall plumbing & TVS	36	0	0	0	0	0	36
Tank wall struts	8	0	0	0	0	0	8
Tank Wall harnesses	8	0	0	0	0	0	8
Panel K inside MLI	12	0	12	0	0	9	33
Panel L inside MLI	16	0	20	0	2	12	50
Struts inside MLI	8	0	0	0	0	0	8
Harnesses inside MLI	8	0	0	0	0	0	8
Honeycomb wall	32	0	0	0	0	0	32
Outside MLI wall	24	0	0	0	0	0	24
Tray & flowmeter	8	0	4	8	0	6	26
Total:	296	22	36	8	2	27	391

TABLE 24.—SMALL RECEIVER MODULE INSTRUMENTATION WIRE ALLOCATION

Location				Sensor type			
	Wires	Liquid- vapor wires	Pressure wires	Differential pressure wires	Turbine meters	Cryogen valve wires	Total wires
Inside pressure vessel	72	22	0	0	0	0	94
Outer tank wall	40	0	0	0	0	0	40
Tank wall plumbing & TVS	32	0	0	0	0	0	32
Tank wall struts	8	0	0	0	0	0	8
Tank wall harnesses	8	0	0	0	0	0	8
Panel M inside MLI	12	0	16	0	0	12	40
Panel N inside MLI	12	0	16	0	0	12	40
Struts inside MLI	8	0	0	0	2	0	10
Harnesses inside MLI	8	0	0	0	0	0	8
Honeycomb wall	40	0	0	0	0	0	40
Outside MLI wall	32	0	0	0	0	0	32
Tray & TVS meter	8	0	4	0	0	6	18
Total:	280	22	36	0	2	30	370

TABLE 25 COLD-SAT INSTRUMENTATION WIRING SUMMARY
(a) Supply tank

Wire location	CTU wires	RCTU 1 wires	RCTU 2 wires	EDU 1	EDU 2	EDU 3	Total
Penetrating tank Thermally bonded to tank Inside MLI Outside MLI & tray Total:	67 68 16 19 170	70 59 16 22 167	48 79 24 19 170	0 16 0 2 18	0 10 0 0	4 8 0 0 12	189 240 56 62 547

(b) Large receiver

Wire location	CTU wires	RCTU 1 wires	RCTU 2 wires	EDU 1	EDU 2	EDU 3	Total
Penetrating tank Thermally bonded to tank Inside MLI Outside MLI & tray Total:	48	46	24	0	0	0	118
	49	42	70	4	4	4	173
	16	16	16	0	0	0	48
	12	19	11	0	6	4	52
	125	123	121	4	10	8	391

(c) Small receiver module

Wire location	CTU wires	RCTU 1 wires	RCTU 2 wires	EDU 1	EDU 2	EDU 3	Total
Penetrating tank Thermally bonded to tank Inside MLI Outside MLI & tray Total:	24 57 20 12 113	24 65 20 12 121	46 38 16 22 122	0 4 0 0 4	0 4 0 0 4	0 6 0 0	94 174 56 46 370

TABLE 26.—COLD-SAT POWER WIRING SOURCE SUMMARY

Wire source destination	Supply tank		Large receiver		Small receiver		Total
	SEQ 1 wires	SEQ 2 wires	SEQ 1 wires	SEQ 2 wires	SEQ 1 wires	SEQ 2 wires	:
Thermally bonded to tank (valves)	36 4	0	20	8 4	20 0	12 4	96 12
Thermally bonded to tank (heater) Outside MLI & tray valves	16	0	0	8	Ö	8	32
Total:	56	0	20	20	20	24	140




Figure 1.—Cryogenic on-orbit liquid depot-storage, acquisition, transfer (COLD-SAT) spacecraft.




Figure 2.—Instrumentation and electronics system and TT&C interface block diagram.




Figure 3.—Simplified data acquisition system block diagram.




Figure 4.—Simplified experiment data unit (EDU) block diagram.




Figure 5.—Data Acquisition system estimated range error for 8-Bit TT&C and 12-Bit EDU systems.




Figure 6.—High-accuracy temperature measurement errors.




Figure 7.—Type AB temperature measurement system error.




Figure 10.—Saturation temperature versus pressure of LH₂.




Figure 8.—Dual range diode temperature measurement system errors. 1.13-V offset (diode) for 32 to 43 °R range.





Figure 9.—1000-Ohm PRT probe and simplified operation schematic.






Figure 11.—Influence on LH $_2$ density of (a) temperature at 30 psia and of (b) pressure at 36 $^{\circ}$ R.




Figure 12.—GH₂ density versus pressure to temperature ratio. Temperature, range, 180 to 360 °R; pressure range, 5 to 20 psia.




Figure 14.—Discrete measurement schematic.




Figure 13.—Rotational speed sensor schematic.





Figure 15.—Thermal conductivity of metals.










Figure 16.—Multiplex use of common excitation wiring to reduce wire penetrations. (a) Temperature sensor wiring. (b) Liquid/vapor sensor multiplexed wire reduction technique.













Figure 18.—Supply tank internal instrumentation. (a) Temperature and liquid vapor sensors. (b) Supply tank temperature sensors.







Figure 19,—COLD-SAT experiment subsystem. Large receiver tank instrumentation locations. (a) Internal instrumentation locations. (b) External temperature sensor locations.






Figure 20.—COLD-SAT experiment subsystem small receiver tank instrumentation locations. (a) Internal instrumentation locations. (b) External temperature sensor locations.

REPORT DOCUMENTATION PAGE

Form Approved OMB No. 0704-0188

Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503.

1. AGENCY USE ONLY (Leave blank)	AGENCY USE ONLY (Leave blank) 2. REPORT DATE 3. REPORT TYPE A				
	March 1994	inal Contractor Report			
4. TITLE AND SUBTITLE	5. FUNDING NUMBERS				
Cryogenic On-Orbit Liquid D					
(COLD-SAT) Experiment Sul					
Wire Harness Design Report					
6. AUTHOR(S)	WU-506-48-00				
			C-NAS3-25776		
Lawrence G. Edwards					
7. PERFORMING ORGANIZATION NA	ME(S) AND ADDRESS(ES)		8. PERFORMING ORGANIZATION		
Analog Composition			REPORT NUMBER		
Analex Corporation 3001 Aerospace Parkway			E-8098		
Brook Park, Ohio 44142			E-8098		
Diook I ark, Ollo 44142					
9. SPONSORING/MONITORING AGEN	CY NAME(S) AND ADDRESS(ES)		10. SPONSORING/MONITORING AGENCY REPORT NUMBER		
National Aeronautics and Spa	ace Administration		AGENO: HEI OH HOMBEN		
Lewis Research Center	uce / tellimistration		NASA CR-189172		
Cleveland, Ohio 44135-319	01				
11. SUPPLEMENTARY NOTES	· · · · · · · · · · · · · · · · · · ·				
Decises Manager James D. Fo	dial Casa Funcionas Dia	!:!!:! 4-	6780 (016) 422 6222		
Project Manager, James R. Fa	ddoui, Space Experiments Div	ision, organization code	6/80, (216) 433-6322.		
12a. DISTRIBUTION/AVAILABILITY ST	12b. DISTRIBUTION CODE				
TT116'-4 TT-1''4-4					
Unclassified - Unlimited	10.20120				
Subject Categories 15, 16, 18,					
13. ABSTRACT (Maximum 200 words)					
	iquid hydrogen (LH ₂) and liqu				
	tant, and life support systems.				
	term cryogen storage and effic	-			
	liquid depot-storage acquisition				
	t. Experiments to determine or		<u> </u>		
	mission. The success of the m				
	rumentation and measurement	_			
	sents the results of the COLD-				
	ate transducers capable of fulfil				
	ignal conditioning techniques,				
	nted. Electrical harnessing mat				
_	nduction to the cryogenic tanks	and provide optimum i	neasurement accuracy are		
listed.					
14. SUBJECT TERMS			15. NUMBER OF PAGES		
COLD-SAT; Cryogenic; Liqui	72				
conditioner; Data acquisition s	16. PRICE CODE				
Manganin; Phosphor-Bronze;	A04				
17. SECURITY CLASSIFICATION 18 OF REPORT	B. SECURITY CLASSIFICATION OF THIS PAGE	19. SECURITY CLASSIFICA OF ABSTRACT	TION 20. LIMITATION OF ABSTRACT		
Unclassified					

11.11.00				
				4.
		-		
ı				
•				
				-
: ▶				
· •				
<u>-</u>				
= - =				